

Data Path : Z:\HPCHEM1\BNA_E\Data\BE020615\
 Data File : BE089541.D
 Acq On : 7 Feb 2015 00:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 07 05:55:29 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
3 S	Nitrobenzene-d5	0.332	0.299	9.9	85	0.00
4	Naphthalene	1.127	1.088	3.5	88	0.00
5	2-Methylnaphthalene	0.640	0.598	6.6	87	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
7 S	2-Fluorobiphenyl	1.334	1.330	0.3	88	0.00
8	Acenaphthylene	8.406	8.029	4.5	85	0.00
9	Acenaphthene	1.335	1.306	2.2	86	0.00
10	Fluorene	1.417	1.298	8.4	85	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
12	Pentachlorophenol	0.030	0.006	80.0#	43#	0.00
13	Phenanthrene	1.239	1.194	3.6	86	0.00
14	Anthracene	1.046	0.980	6.3	85	0.00
15	Fluoranthene	1.154	1.066	7.6	85	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
17	Pyrene	1.065	0.952	10.6	88	0.00
18 S	Terphenyl-d14	0.660	0.578	12.4	85	0.00
19	Benzo(a)anthracene	4.449	3.824	14.0	86	0.00
20	Chrysene	4.519	4.514	0.1	95	0.00
21	Indeno(1,2,3-cd)pyrene	6.282	5.676	9.6	80	0.00
22 I	Perylene-d12	1.000	1.000	0.0	92	0.00
23	Benzo(b)fluoranthene	0.971	0.767	21.0	79	0.00
24	Benzo(k)fluoranthene	1.207	1.141	5.5	91	0.00
25 C	Benzo(a)pyrene	0.966	0.788	18.4	79	0.00
26	Dibenzo(a,h)anthracene	1.185	1.050	11.4	80	0.00
27	Benzo(g,h,i)perylene	4.838	4.466	7.7	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0