

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020618\
 Data File : BE095538.D
 Acq On : 6 Feb 2018 20:27
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE020618

Quant Time: Feb 06 21:03:20 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 06 20:27:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.662	0.658	0.6	99	0.00
3	n-Nitrosodimethylamine	0.833	0.818	1.8	97	0.00
4 S	2-Fluorophenol	1.207	1.179	2.3	96	0.00
5 S	Phenol-d6	1.671	1.590	4.8	94	0.00
6	bis(2-Chloroethyl)ether	1.538	1.546	-0.5	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.424	0.412	2.8	94	0.00
9	Naphthalene	4.646	4.738	-2.0	97	0.00
10	Hexachlorobutadiene	0.249	0.270	-8.4	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.662	0.672	-1.5	97	0.00
12	2-Methylnaphthalene	0.789	0.801	-1.5	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.182	0.155	14.8	86	0.00
15 S	2-Fluorobiphenyl	1.725	1.733	-0.5	98	0.00
16	Acenaphthylene	2.282	2.195	3.8	95	0.00
17	Acenaphthene	1.418	1.391	1.9	97	0.00
18	Fluorene	1.782	1.740	2.4	95	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.240	0.229	4.6	92	0.00
21	Hexachlorobenzene	0.248	0.249	-0.4	98	0.00
22	Pentachlorophenol	0.069	0.052	24.6	90	0.00
23	Phenanthrene	1.228	1.220	0.7	97	-0.01
24	Anthracene	1.147	1.106	3.6	96	-0.01
25 SURR	Fluoranthene-d10	5.268	5.139	2.4	97	0.00
26	Fluoranthene	1.550	1.529	1.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.01
28	Pvrene	1.680	1.641	2.3	104	0.00
29 S	Terphenyl-d14	0.852	0.838	1.6	98	0.00
30	Benzo(a)anthracene	1.333	1.268	4.9	96	0.00
31	Chrysene	1.314	1.283	2.4	97	0.00
32	Bis(2-ethylhexyl)phthalate	3.072	2.527	17.7	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.269	1.190	6.2	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.364	1.377	-1.0	97	0.00
36	Benzo(k)fluoranthene	1.402	1.423	-1.5	98	0.00
37 C	Benzo(a)pyrene	1.269	1.250	1.5	96	0.00
38	Dibenzo(a,h)anthracene	1.111	1.090	1.9	98	0.00
39	Benzo(a,h,i)perylene	1.193	1.167	2.2	95	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020618\
Data File : BE095538.D
Acq On : 6 Feb 2018 20:27
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE020618

Quant Time: Feb 06 21:03:20 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 20:27:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			