

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
 Data File : BE095572.D
 Acq On : 8 Feb 2018 16:55
 Operator : SJ/JU
 Sample : J1315-20
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C83

Manual Integrations
 APPROVED

Sohil
 2/9/2018 4:30:50 PM

Quant Time: Feb 09 03:13:48 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 09 01:31:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1960	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	5891	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	283492	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	8435m	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	6815	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	6306m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	3633	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	10166	0.31	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.30	128	6470	0.392	ng/ul#	1
8) Acenaphthene	15.08	153	108008	0.102	ng/ul	93
9) Fluorene	16.02	166	116879m	0.107	ng/ul	
11) Pentachlorophenol	17.36	266	3762	2.303	ng/ul	96
12) Phenanthrene	17.75	178	16974m	0.598	ng/ul	
13) Anthracene	17.84	178	81393	3.135	ng/ul#	92
15) Fluoranthene	19.72	202	302349	8.231	ng/ul	82
17) Pyrene	20.08	202	238334m	10.422	ng/ul	
18) Benzo(a)anthracene	21.78	228	45278	2.125	ng/ul#	83
19) Chrysene	21.84	228	38740	1.751	ng/ul	92
21) Benzo(b)fluoranthene	23.62	252	29699m	1.335	ng/ul	
22) Benzo(k)fluoranthene	23.66	252	8120m	0.399	ng/ul	
23) Benzo(a)pyrene	24.32	252	16142m	0.813	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.23	276	5037m	0.231	ng/ul	
25) Dibenzo(a,h)anthracene	27.25	278	1457m	0.082	ng/ul	
26) Benzo(a,h,i)perylene	28.11	276	4048m	0.219	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
Data File : BE095572.D
Acq On : 8 Feb 2018 16:55
Operator : SJ/JU
Sample : J1315-20
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6C83

Manual Integrations
APPROVED

Sohil
2/9/2018 4:30:50 PM

Quant Time: Feb 09 03:13:48 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 09 01:31:22 2018
Response via : Initial Calibration

