

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
 Data File : BE095582.D
 Acq On : 8 Feb 2018 22:53
 Operator : SJ/JU
 Sample : J1316-08
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C91

Manual Integrations
 APPROVED

Sohil
 2/9/2018 4:31:20 PM

Quant Time: Feb 09 04:01:25 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 09 01:31:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1718	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	4847	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	3149	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	7253	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	7251	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	6577m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	2312	0.29	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	6553	0.24	ng/ul	0.00
Target Compounds				Ovalue		
11) Pentachlorophenol	17.36	266	1822	1.297	ng/ul	98
12) Phenanthrene	17.75	178	1736m	0.071	ng/ul	
13) Anthracene	17.85	178	485m	0.022	ng/ul	
15) Fluoranthene	19.72	202	2027	0.064	ng/ul	85
17) Pyrene	20.07	202	1744m	0.072	ng/ul	
18) Benzo(a)anthracene	21.78	228	495	0.022	ng/ul#	78
19) Chrysene	21.84	228	582	0.025	ng/ul#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
Data File : BE095582.D
Acq On : 8 Feb 2018 22:53
Operator : SJ/JU
Sample : J1316-08
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6C91

Manual Integrations
APPROVED

Sohil
2/9/2018 4:31:20 PM

Quant Time: Feb 09 04:01:25 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 09 01:31:22 2018
Response via : Initial Calibration

