

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
 Data File : BE095561.D
 Acq On : 8 Feb 2018 10:21
 Operator : SJ/JU
 Sample : J1296-04DL 5X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C22DL

Quant Time: Feb 09 00:51:48 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 09 00:43:28 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1542	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	3991	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	2314	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	5961	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	6387	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	6232	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	469	0.07	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	1363	0.06	ng/ul	0.00
Target Compounds						
3) Naphthalene	11.30	128	535	0.048	ng/ul#	66
11) Pentachlorophenol	17.36	266	1040	0.901	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020818\
Data File : BE095561.D
Acq On : 8 Feb 2018 10:21
Operator : SJ/JU
Sample : J1296-04DL 5X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6C22DL

Quant Time: Feb 09 00:51:48 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 09 00:43:28 2018
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.048 ng/ul

RT: 11.30 min Scan# 1347

Delta R.T. -0.00 min

Lab File: BE095561.D

Acq: 8 Feb 2018 10:21

Instrument :

BNA_E

ClientSampleId :

F6C22DL


Tot Ion:128 Resp: 535

Ion Ratio Lower Upper


128 100

129 23.9 11.3 16.9#



127 25.1 4.0 6.0#



Abundance Ion 128.00 (127.70 to 128.70): E



Time-->



#11

Pentachlorophenol

Concen: 0.901 ng/ul

RT: 17.36 min Scan# 1997

Delta R.T. 0.00 min

Lab File: BE095561.D

Acq: 8 Feb 2018 10:21


Tgt Ion:266 Resp: 1040

Ion Ratio Lower Upper


266 100

264 63.0 47.9 71.9

268 63.9 49.8 74.8



Abundance Ion 266.00 (265.70 to 266.70): E



Time-->

