

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020918\
 Data File : BE095637.D
 Acq On : 10 Feb 2018 8:54
 Operator : SJ/JU
 Sample : J1355-08
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6D52

Manual Integrations
 APPROVED

Sohil
 2/12/2018 11:40:24 AM

Quant Time: Feb 12 01:20:26 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 10 02:59:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	2039	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	6299	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	4300	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	9879	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	8719	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	8050m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	4559	0.44	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	12824	0.34	ng/ul	0.00
Target Compounds						
11) Pentachlorophenol	17.36	266	322	0.168	ng/ul	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020918\
Data File : BE095637.D
Acq On : 10 Feb 2018 8:54
Operator : SJ/JU
Sample : J1355-08
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6D52

Manual Integrations
APPROVED

Sohil
2/12/2018 11:40:24 AM

Quant Time: Feb 12 01:20:26 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 10 02:59:57 2018
Response via : Initial Calibration

