

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020918\
 Data File : BE095645.D
 Acq On : 10 Feb 2018 14:17
 Operator : SJ/JU
 Sample : J1355-04
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6D29

Manual Integrations
 APPROVED

Sohil
 2/12/2018 11:40:44 AM

Quant Time: Feb 12 02:41:16 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 12 01:48:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	2127	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	6549	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	4447	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	10013	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	9035	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	8458m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	4450	0.42	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	12134	0.32	ng/ul	0.00
Target Compounds						
11) Pentachlorophenol	17.36	266	133	0.069	ng/ul	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020918\
Data File : BE095645.D
Acq On : 10 Feb 2018 14:17
Operator : SJ/JU
Sample : J1355-04
Misc :
ALS Vial : 45 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
F6D29

Manual Integrations
APPROVED

Sohil
2/12/2018 11:40:44 AM


Quant Time: Feb 12 02:41:16 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 12 01:48:13 2018
Response via : Initial Calibration





#11
Pentachlorophenol
Concen: 0.069 ng/ul
RT: 17.36 min Scan# 1997
Delta R.T. 0.00 min
Lab File: BE095645.D
Acq: 10 Feb 2018 14:17

Instrument :
BNA_E
ClientSampleId :
F6D29



Tgt Ion:	266	Resp:	133
Ion Ratio	Lower	Upper	
266	100		
264	56.4	47.9	71.9
268	63.2	49.8	74.8

Manual Integrations
APPROVED

Sohil
2/12/2018 11:40:44 AM

