

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021015\
 Data File : BE089552.D
 Acq On : 10 Feb 2015 16:17
 Operator : TP/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.1

Quant Time: Feb 11 01:21:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 01:08:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.42	152	34437	5.00	ng	0.00
2) Naphthalene-d8	12.32	136	131145	5.00	ng	0.00
6) Acenaphthene-d10	16.50	164	54333	5.00	ng	0.00
11) Phenanthrene-d10	19.84	188	108830	5.00	ng	0.00
16) Chrysene-d12	23.71	240	129591	5.00	ng	0.00
22) Perylene-d12	25.40	264	130755	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.68	82	715	0.09	ng	0.00
7) 2-Fluorobiphenyl	14.95	172	1420	0.11	ng	0.00
18) Terphenyl-d14	22.36	244	1578	0.09	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.37	128	3016	0.11	ng	# 93
5) 2-Methylnaphthalene	14.03	142	1597	0.10	ng	97
8) Acenaphthylene	16.15	152	8389	0.10	ng	98
9) Acenaphthene	16.58	154	1450	0.11	ng	88
10) Fluorene	17.83	166	1583	0.11	ng	# 99
12) Pentachlorophenol	19.51	266	187	0.06	ng	91
13) Phenanthrene	19.88	178	2651m	0.11	ng	98
14) Anthracene	19.97	178	1922	0.09	ng	99
15) Fluoranthene	21.78	202	2292	0.10	ng	100
17) Pyrene	22.10	202	2456	0.09	ng	99
19) Benzo(a)anthracene	23.70	228	10489	0.10	ng	98
20) Chrysene	23.74	228	12301	0.11	ng	86
21) Indeno(1,2,3-cd)pyrene	26.83	276	7833	0.08	ng	# 92
23) Benzo(b)fluoranthene	24.96	252	1594	0.08	ng	# 86
24) Benzo(k)fluoranthene	25.00	252	2525m	0.08	ng	# 85
25) Benzo(a)pyrene	25.33	252	1502	0.08	ng	# 89
26) Dibenzo(a,h)anthracene	26.86	278	1544	0.08	ng	
27) Benzo(g,h,i)perylene	27.26	276	8048	0.09	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021015\
Data File : BE089552.D
Acq On : 10 Feb 2015 16:17
Operator : TP/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.1

Quant Time: Feb 11 01:21:38 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 11 01:08:43 2015
Response via : Initial Calibration

