

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021015\
 Data File : BE089560.D
 Acq On : 10 Feb 2015 21:43
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICV001

Quant Time: Feb 11 11:33:20 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 11:21:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.41	152	25462	5.00	ng	0.00
2) Naphthalene-d8	12.32	136	90423	5.00	ng	0.00
6) Acenaphthene-d10	16.50	164	30098	5.00	ng	0.00
11) Phenanthrene-d10	19.84	188	45547	5.00	ng	0.00
16) Chrysene-d12	23.71	240	58735	5.00	ng	0.00
22) Perylene-d12	25.40	264	71317	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.68	82	5963	1.03	ng	0.00
7) 2-Fluorobiphenyl	14.95	172	8781	1.08	ng	0.00
18) Terphenyl-d14	22.36	244	6489	0.89	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.37	128	20492	1.01	ng	99
5) 2-Methylnaphthalene	14.03	142	10906	0.98	ng	99
8) Acenaphthylene	16.15	152	52485	1.02	ng	100
9) Acenaphthene	16.58	154	8220	1.02	ng	98
10) Fluorene	17.83	166	7973	0.93	ng	98
12) Pentachlorophenol	19.51	266	218	0.86	ng	# 87
13) Phenanthrene	19.88	178	11436	1.03	ng	100
14) Anthracene	19.97	178	9062	0.96	ng	98
15) Fluoranthene	21.78	202	9738	0.91	ng	99
17) Pyrene	22.10	202	10925	0.92	ng	99
19) Benzo(a)anthracene	23.70	228	44216	0.90	ng	99
20) Chrysene	23.74	228	56547	1.03	ng	100
21) Indeno(1,2,3-cd)pyrene	26.83	276	65583	0.99	ng	# 80
23) Benzo(b)fluoranthene	24.97	252	8788	0.82	ng	99
24) Benzo(k)fluoranthene	25.00	252	18216	1.04	ng	99
25) Benzo(a)pyrene	25.34	252	9846	0.84	ng	98
26) Dibenzo(a,h)anthracene	26.87	278	12904	0.87	ng	98
27) Benzo(g,h,i)perylene	27.26	276	62971	1.07	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021015\
Data File : BE089560.D
Acq On : 10 Feb 2015 21:43
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICV001

Quant Time: Feb 11 11:33:20 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 11 11:21:59 2015
Response via : Initial Calibration

