

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021015\
 Data File : BE089560.D
 Acq On : 10 Feb 2015 21:43
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICV001

Quant Time: Feb 11 11:33:20 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 11:21:59 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
3 S	Nitrobenzene-d5	0.320	0.330	-3.1	110	0.00
4	Naphthalene	1.119	1.133	-1.3	103	0.00
5	2-Methylnaphthalene	0.618	0.603	2.4	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
7 S	2-Fluorobiphenyl	1.355	1.459	-7.7	101	0.00
8	Acenaphthylene	8.523	8.719	-2.3	101	0.00
9	Acenaphthene	1.337	1.366	-2.2	101	0.00
10	Fluorene	1.430	1.325	7.3	98	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
12	Pentachlorophenol	0.043	0.024	44.2#	95	0.00
13	Phenanthrene	1.225	1.255	-2.4	95	0.00
14	Anthracene	1.037	0.995	4.1	92	0.00
15	Fluoranthene	1.181	1.069	9.5	90	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
17	Pyrene	1.006	0.930	7.6	91	0.00
18 S	Terphenyl-d14	0.623	0.552	11.4	85	0.00
19	Benzo(a)anthracene	4.160	3.764	9.5	87	0.00
20	Chrysene	4.680	4.814	-2.9	93	0.00
21	Indeno(1,2,3-cd)pyrene	4.951	5.583	-12.8	93	0.00
22 I	Perylene-d12	1.000	1.000	0.0	94	0.00
23	Benzo(b)fluoranthene	0.773	0.616	20.3	84	0.00
24	Benzo(k)fluoranthene	1.226	1.277	-4.2	94	0.00
25 C	Benzo(a)pyrene	0.797	0.690	13.4	88	0.00
26	Dibenzo(a,h)anthracene	0.946	0.905	4.3	88	0.00
27	Benzo(g,h,i)perylene	4.118	4.415	-7.2	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0