

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021015\
 Data File : BE089567.D
 Acq On : 11 Feb 2015 2:56
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc : MDL-WATER-0.1/0.2
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 11 03:51:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 01:27:13 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
3 S	Nitrobenzene-d5	0.320	0.310	3.1	111	0.00
4	Naphthalene	1.119	1.127	-0.7	110	0.00
5	2-Methylnaphthalene	0.618	0.624	-1.0	113	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
7 S	2-Fluorobiphenyl	1.355	1.339	1.2	117	0.00
8	Acenaphthylene	8.523	8.400	1.4	124	0.00
9	Acenaphthene	1.337	1.327	0.7	125	0.00
10	Fluorene	1.430	1.435	-0.3	135	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	148	0.00
12	Pentachlorophenol	0.043	0.024	44.2#	155#	0.00
13	Phenanthrene	1.225	1.224	0.1	149	0.00
14	Anthracene	1.037	1.017	1.9	153#	0.00
15	Fluoranthene	1.181	1.153	2.4	156#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	140	0.00
17	Pyrene	1.006	1.033	-2.7	157#	0.00
18 S	Terphenyl-d14	0.623	0.625	-0.3	149	0.00
19	Benzo(a)anthracene	4.160	3.718	10.6	133	0.00
20	Chrysene	4.680	4.813	-2.8	143	0.00
21	Indeno(1,2,3-cd)pyrene	4.951	3.582	27.7#	92	0.00
22 I	Perylene-d12	1.000	1.000	0.0	121	0.00
23	Benzo(b)fluoranthene	0.773	0.516	33.2#	90	0.00
24	Benzo(k)fluoranthene	1.226	1.317	-7.4	126	0.00
25 C	Benzo(a)pyrene	0.797	0.617	22.6#	102	0.00
26	Dibenzo(a,h)anthracene	0.946	0.695	26.5#	87	0.00
27	Benzo(g,h,i)perylene	4.118	3.600	12.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1