

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021015\
 Data File : BE089567.D
 Acq On : 11 Feb 2015 2:56
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc : MDL-WATER-0.1/0.2
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 11 03:51:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 11 01:27:13 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	106	0.00
2 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
3 S	Nitrobenzene-d5	1.000	0.969	3.1	111	0.00
4	Naphthalene	1.000	1.007	-0.7	110	0.00
5	2-Methylnaphthalene	1.000	1.010	-1.0	113	0.00
6 I	Acenaphthene-d10	5.000	5.000	0.0	124	0.00
7 S	2-Fluorobiphenyl	1.000	0.988	1.2	117	0.00
8	Acenaphthylene	1.000	0.986	1.4	124	0.00
9	Acenaphthene	1.000	0.992	0.8	125	0.00
10	Fluorene	1.000	1.004	-0.4	135	0.00
11 I	Phenanthrene-d10	5.000	5.000	0.0	148	0.00
12	Pentachlorophenol	1.000	0.916	8.4	155	0.00
13	Phenanthrene	1.000	0.999	0.1	149	0.00
14	Anthracene	1.000	0.981	1.9	153	0.00
15	Fluoranthene	1.000	0.976	2.4	156	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	140	0.00
17	Pyrene	1.000	1.026	-2.6	157	0.00
18 S	Terphenyl-d14	1.000	1.003	-0.3	149	0.00
19	Benzo(a)anthracene	1.000	0.894	10.6	133	0.00
20	Chrysene	1.000	1.028	-2.8	143	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	0.651	34.9#	92	0.00
22 I	Perylene-d12	5.000	5.000	0.0	121	0.00
23	Benzo(b)fluoranthene	1.000	0.727	27.3#	90	0.00
24	Benzo(k)fluoranthene	1.000	1.075	-7.5	126	0.00
25 C	Benzo(a)pyrene	1.000	0.771	22.9#	102	0.00
26	Dibenzo(a,h)anthracene	1.000	0.700	30.0#	87	0.00
27	Benzo(g,h,i)perylene	1.000	0.874	12.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1