

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021017\
 Data File : BE092722.D
 Acq On : 10 Feb 2017 15:29
 Operator : SJ/MA
 Sample : I1744-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS027-170206

Manual Integrations
 APPROVED

mohammad
 2/13/2017 12:51:06 PM

Quant Time: Feb 10 16:14:28 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	10218	5.00	ng	-0.02
7) Naphthalene-d8	10.89	136	43421	5.00	ng	-0.02
12) Acenaphthene-d10	14.77	164	27400	5.00	ng	-0.02
18) Phenanthrene-d10	17.53	188	52866	5.00	ng	0.00
24) Chrysene-d12	21.70	240	80270	5.00	ng	-0.02
31) Perylene-d12	24.04	264	72275	5.00	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.62	112	8823	3.87	ng	-0.03
5) Phenol-d6	7.29	99	6283	2.25	ng	-0.05
8) Nitrobenzene-d5	9.29	82	15322	5.44	ng	-0.05
13) 2,4,6-Tribromophenol	16.26	330	8353	10.29	ng	-0.04
14) 2-Fluorobiphenyl	13.39	172	37102	5.51	ng	-0.03
26) Terphenyl-d14	20.14	244	51091	4.50	ng	-0.02
Target Compounds						
21) Phenanthrene	17.56	178	524m	0.04	ng	Qvalue
22) Anthracene	17.67	178	771	0.07	ng	# 93
23) Fluoranthene	19.56	202	3971	0.08	ng	98
25) Pyrene	19.93	202	3740	0.05	ng	97
27) Benzo(a)anthracene	21.68	228	2858m	0.04	ng	
28) Chrysene	21.72	228	3415m	0.04	ng	
29) Bis(2-ethylhexyl)phthalate	21.59	149	2219	0.34	ng	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021017\
Data File : BE092722.D
Acq On : 10 Feb 2017 15:29
Operator : SJ/MA
Sample : I1744-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS027-170206

Manual Integrations
APPROVED

mohammad
2/13/2017 12:51:06 PM

Quant Time: Feb 10 16:14:28 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 10:14:01 2017
Response via : Initial Calibration

