

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021017\
 Data File : BE092718.D
 Acq On : 10 Feb 2017 12:21
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 10 14:46:07 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.02
2	1,4-Dioxane	0.775	0.731	5.7	88	0.00
3	n-Nitrosodimethylamine	0.902	0.668	25.9#	83	0.00
4 S	2-Fluorophenol	1.116	1.124	-0.7	116	-0.01
5 S	Phenol-d6	1.368	1.447	-5.8	126	-0.02
6	bis(2-Chloroethyl)ether	1.453	1.199	17.5	96	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	93	-0.02
8 S	Nitrobenzene-d5	0.325	0.320	1.5	111	-0.02
9	Naphthalene	1.044	1.123	-7.6	108	-0.02
10	Hexachlorobutadiene	0.154	0.170	-10.4	108	-0.02
11	2-Methylnaphthalene	0.648	0.668	-3.1	107	-0.02
12 I	Acenaphthene-d10	1.000	1.000	0.0	91	-0.02
13 S	2,4,6-Tribromophenol	0.126	0.135	-7.1	123	-0.01
14 S	2-Fluorobiphenyl	1.230	1.322	-7.5	106	-0.02
15	Acenaphthylene	6.890	7.337	-6.5	110	-0.02
16	Acenaphthene	1.105	1.191	-7.8	104	-0.02
17	Fluorene	1.387	1.423	-2.6	105	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	81	0.00
19	Hexachlorobenzene	0.189	0.215	-13.8	88	-0.02
20	Pentachlorophenol	0.051	0.050	2.0	86	0.05
21	Phenanthrene	1.104	1.247	-13.0	95	-0.02
22	Anthracene	1.018	1.124	-10.4	100	-0.02
23	Fluoranthene	4.895	5.794	-18.4	106	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	95	-0.02
25	Pyrene	4.929	4.395	10.8	102	-0.02
26 S	Terphenyl-d14	0.707	0.732	-3.5	115	-0.02
27	Benzo(a)anthracene	4.926	4.990	-1.3	112	-0.02
28	Chrysene	5.542	5.126	7.5	95	-0.02
29	Bis(2-ethylhexyl)phthalate	0.913	0.655	28.3#	96	-0.02
30	Indeno(1,2,3-cd)pyrene	5.987	5.691	4.9	104	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	83	-0.02
32	Benzo(b)fluoranthene	1.210	1.282	-6.0	96	-0.02
33	Benzo(k)fluoranthene	1.215	1.274	-4.9	101	-0.02
34 C	Benzo(a)pyrene	1.136	1.202	-5.8	102	-0.02
35	Dibenzo(a,h)anthracene	1.112	1.235	-11.1	104	-0.05
36	Benzo(g,h,i)perylene	4.699	5.270	-12.2	103	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0