

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021017\
 Data File : BE092726.D
 Acq On : 10 Feb 2017 17:58
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 10 18:35:34 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	-0.02
2	1,4-Dioxane	0.775	0.746	3.7	88	0.00
3	n-Nitrosodimethylamine	0.902	0.607	32.7#	74	0.01
4 S	2-Fluorophenol	1.116	0.962	13.8	98	0.00
5 S	Phenol-d6	1.368	1.260	7.9	108	-0.02
6	bis(2-Chloroethyl)ether	1.453	1.139	21.6	90	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	-0.02
8 S	Nitrobenzene-d5	0.325	0.286	12.0	102	-0.02
9	Naphthalene	1.044	1.090	-4.4	109	-0.02
10	Hexachlorobutadiene	0.154	0.170	-10.4	112	-0.02
11	2-Methylnaphthalene	0.648	0.653	-0.8	108	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.02
13 S	2,4,6-Tribromophenol	0.126	0.107	15.1	100	0.00
14 S	2-Fluorobiphenyl	1.230	1.306	-6.2	107	-0.02
15	Acenaphthylene	6.890	7.120	-3.3	109	-0.02
16	Acenaphthene	1.105	1.185	-7.2	106	-0.02
17	Fluorene	1.387	1.366	1.5	103	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
19	Hexachlorobenzene	0.189	0.230	-21.7	96	-0.02
20	Pentachlorophenol	0.051	0.035	31.4#	61	0.07
21	Phenanthrene	1.104	1.303	-18.0	100	-0.02
22	Anthracene	1.018	1.132	-11.2	102	-0.02
23	Fluoranthene	4.895	6.091	-24.4	112	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	101	-0.02
25	Pyrene	4.929	4.563	7.4	112	-0.02
26 S	Terphenyl-d14	0.707	0.656	7.2	109	-0.02
27	Benzo(a)anthracene	4.926	4.616	6.3	110	-0.02
28	Chrysene	5.542	4.981	10.1	97	-0.02
29	Bis(2-ethylhexyl)phthalate	0.913	0.403	55.9#	62	-0.02
30	Indeno(1,2,3-cd)pyrene	5.987	5.019	16.2	97	-0.03
31 I	Perylene-d12	1.000	1.000	0.0	80	-0.02
32	Benzo(b)fluoranthene	1.210	1.280	-5.8	92	-0.02
33	Benzo(k)fluoranthene	1.215	1.295	-6.6	99	-0.02
34 C	Benzo(a)pyrene	1.136	1.173	-3.3	95	-0.02
35	Dibenzo(a,h)anthracene	1.112	1.184	-6.5	96	-0.05
36	Benzo(g,h,i)perylene	4.699	5.093	-8.4	96	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0