

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021017\
 Data File : BE092726.D
 Acq On : 10 Feb 2017 17:58
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 10 18:35:34 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	-0.02
2	1,4-Dioxane	0.400	0.379	5.3	88	0.00
3	n-Nitrosodimethylamine	0.400	0.269	32.8#	74	0.01
4 S	2-Fluorophenol	0.400	0.345	13.8	98	0.00
5 S	Phenol-d6	0.400	0.368	8.0	108	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.314	21.5	90	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	96	-0.02
8 S	Nitrobenzene-d5	0.400	0.352	12.0	102	-0.02
9	Naphthalene	0.400	0.417	-4.2	109	-0.02
10	Hexachlorobutadiene	0.400	0.442	-10.5	112	-0.02
11	2-Methylnaphthalene	0.400	0.403	-0.8	108	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	94	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.415	-3.7	100	0.00
14 S	2-Fluorobiphenyl	0.400	0.425	-6.2	107	-0.02
15	Acenaphthylene	0.400	0.413	-3.2	109	-0.02
16	Acenaphthene	0.400	0.429	-7.2	106	-0.02
17	Fluorene	0.400	0.394	1.5	103	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	82	0.00
19	Hexachlorobenzene	0.400	0.481	-20.2	96	-0.02
20	Pentachlorophenol	0.400	0.274	31.5#	61	0.07
21	Phenanthrene	0.400	0.472	-18.0	100	-0.02
22	Anthracene	0.400	0.445	-11.2	102	-0.02
23	Fluoranthene	0.400	0.498	-24.5	112	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	101	-0.02
25	Pyrene	0.400	0.370	7.5	112	-0.02
26 S	Terphenyl-d14	0.400	0.371	7.3	109	-0.02
27	Benzo(a)anthracene	0.400	0.375	6.3	110	-0.02
28	Chrysene	0.400	0.360	10.0	97	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.365	8.8	62	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.335	16.3	97	-0.03
31 I	Perylene-d12	5.000	5.000	0.0	80	-0.02
32	Benzo(b)fluoranthene	0.400	0.423	-5.7	92	-0.02
33	Benzo(k)fluoranthene	0.400	0.427	-6.7	99	-0.02
34 C	Benzo(a)pyrene	0.400	0.413	-3.2	95	-0.02
35	Dibenzo(a,h)anthracene	0.400	0.426	-6.5	96	-0.05
36	Benzo(g,h,i)perylene	0.400	0.434	-8.5	96	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0