

Data Path : Z:\HPCHEM1\BNA E\Data\BE021318\
 Data File : BE095656.D
 Acq On : 12 Feb 2018 11:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 13 03:23:15 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 06 20:27:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.662	0.679	-2.6	98	0.00
3	n-Nitrosodimethylamine	0.833	0.881	-5.8	100	-0.01
4 S	2-Fluorophenol	1.207	1.199	0.7	94	0.00
5 S	Phenol-d6	1.671	1.663	0.5	94	-0.01
6	bis(2-Chloroethyl)ether	1.538	1.570	-2.1	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.02
8 S	Nitrobenzene-d5	0.424	0.398	6.1	93	-0.02
9	Naphthalene	4.646	4.695	-1.1	98	-0.02
10	Hexachlorobutadiene	0.249	0.264	-6.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.662	0.697	-5.3	103	0.00
12	2-Methylnaphthalene	0.789	0.821	-4.1	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.182	0.152	16.5	89	0.00
15 S	2-Fluorobiphenyl	1.725	1.744	-1.1	104	-0.02
16	Acenaphthylene	2.282	2.236	2.0	102	0.00
17	Acenaphthene	1.418	1.381	2.6	101	0.00
18	Fluorene	1.782	1.738	2.5	100	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.240	0.250	-4.2	98	-0.01
21	Hexachlorobenzene	0.248	0.259	-4.4	100	0.00
22	Pentachlorophenol	0.069	0.063	8.7	106	0.00
23	Phenanthrene	1.228	1.254	-2.1	98	-0.01
24	Anthracene	1.147	1.154	-0.6	99	-0.01
25 SURR	Fluoranthene-d10	5.268	5.144	2.4	96	0.00
26	Fluoranthene	1.550	1.517	2.1	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.680	1.759	-4.7	105	0.00
29 S	Terphenyl-d14	0.852	0.839	1.5	92	0.00
30	Benzo(a)anthracene	1.333	1.387	-4.1	99	0.00
31	Chrysene	1.314	1.347	-2.5	96	0.00
32	Bis(2-ethylhexyl)phthalate	3.072	2.670	13.1	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.269	1.370	-8.0	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.364	1.374	-0.7	97	0.00
36	Benzo(k)fluoranthene	1.402	1.417	-1.1	98	0.00
37 C	Benzo(a)pyrene	1.269	1.292	-1.8	100	0.00
38	Dibenzo(a,h)anthracene	1.111	1.195	-7.6	108	0.00
39	Benzo(a,h,i)perylene	1.193	1.235	-3.5	101	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE021318\
Data File : BE095656.D
Acq On : 12 Feb 2018 11:55
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 13 03:23:15 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 20:27:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			