

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021517\
 Data File : BE092736.D
 Acq On : 15 Feb 2017 16:52
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021517

Quant Time: Feb 15 17:30:31 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	1,4-Dioxane	0.400	0.383	4.3	96	0.00
3	n-Nitrosodimethylamine	0.400	0.459	-14.8	94	0.01
4 S	2-Fluorophenol	0.400	0.362	9.5	95	0.00
5 S	Phenol-d6	0.400	0.371	7.3	97	0.00
6	bis(2-Chloroethyl)ether	0.400	0.366	8.5	100	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	97	0.00
9	Naphthalene	0.400	0.402	-0.5	100	0.00
10	Hexachlorobutadiene	0.400	0.414	-3.5	100	0.00
11	2-Methylnaphthalene	0.400	0.394	1.5	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.400	0.351	12.3	93	0.00
14 S	2-Fluorobiphenyl	0.400	0.396	1.0	101	0.00
15	Acenaphthylene	0.400	0.381	4.8	100	0.00
16	Acenaphthene	0.400	0.397	0.8	102	0.00
17	Fluorene	0.400	0.384	4.0	101	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.00
19	Hexachlorobenzene	0.400	0.428	-7.0	99	0.00
20	Pentachlorophenol	0.400	0.478	-19.5	93	0.00
21	Phenanthrene	0.400	0.402	-0.5	97	0.00
22	Anthracene	0.400	0.375	6.3	98	0.00
23	Fluoranthene	0.400	0.390	2.5	99	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
25	Pyrene	0.400	0.388	3.0	102	0.00
26 S	Terphenyl-d14	0.400	0.383	4.3	101	0.00
27	Benzo(a)anthracene	0.400	0.390	2.5	102	0.00
28	Chrysene	0.400	0.393	1.8	100	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.377	5.8	95	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.418	-4.5	91	0.00
31 I	Perylene-d12	5.000	5.000	0.0	100	0.00
32	Benzo(b)fluoranthene	0.400	0.427	-6.7	95	0.00
33	Benzo(k)fluoranthene	0.400	0.383	4.3	97	0.00
34 C	Benzo(a)pyrene	0.400	0.364	9.0	97	0.00
35	Dibenzo(a,h)anthracene	0.400	0.392	2.0	91	0.00
36	Benzo(g,h,i)perylene	0.400	0.396	1.0	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0