

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021517\
 Data File : BE092736.D
 Acq On : 15 Feb 2017 16:52
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021517

Quant Time: Feb 15 17:30:31 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.710	0.667	6.1	96	0.00
3	n-Nitrosodimethylamine	0.740	0.601	18.8	94	0.01
4 S	2-Fluorophenol	0.973	0.880	9.6	95	0.00
5 S	Phenol-d6	1.194	1.106	7.4	97	0.00
6	bis(2-Chloroethyl)ether	1.321	1.207	8.6	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.279	0.260	6.8	97	0.00
9	Naphthalene	1.082	1.088	-0.6	100	0.00
10	Hexachlorobutadiene	0.159	0.165	-3.8	100	0.00
11	2-Methylnaphthalene	0.680	0.670	1.5	102	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.106	0.093	12.3	93	0.00
14 S	2-Fluorobiphenyl	1.219	1.208	0.9	101	0.00
15	Acenaphthylene	7.544	7.193	4.7	100	0.00
16	Acenaphthene	1.096	1.087	0.8	102	0.00
17	Fluorene	1.407	1.349	4.1	101	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
19	Hexachlorobenzene	0.193	0.207	-7.3	99	0.00
20	Pentachlorophenol	0.042	0.033	21.4	93	0.00
21	Phenanthrene	1.188	1.195	-0.6	97	0.00
22	Anthracene	1.094	1.027	6.1	98	0.00
23	Fluoranthene	5.096	4.970	2.5	99	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
25	Pyrene	4.761	4.615	3.1	102	0.00
26 S	Terphenyl-d14	0.638	0.612	4.1	101	0.00
27	Benzo(a)anthracene	4.849	4.726	2.5	102	0.00
28	Chrysene	5.752	5.645	1.9	100	0.00
29	Bis(2-ethylhexyl)phthalate	0.674	0.538	20.2	95	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	5.821	10.3	91	0.00
31 I	Perylene-d12	1.000	1.000	0.0	100	0.00
32	Benzo(b)fluoranthene	1.414	1.305	7.7	95	0.00
33	Benzo(k)fluoranthene	1.495	1.327	11.2	97	0.00
34 C	Benzo(a)pyrene	1.285	1.170	8.9	97	0.00
35	Dibenzo(a,h)anthracene	1.355	1.218	10.1	91	0.00
36	Benzo(g,h,i)perylene	5.728	5.148	10.1	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0