

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021615\
 Data File : BE089570.D
 Acq On : 16 Feb 2015 15:03
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC005

Quant Time: Feb 16 17:17:48 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIM-BE021615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 16 17:05:37 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	82021	5.00	ng	0.00
7) Naphthalene-d8	12.25	136	327815	5.00	ng	0.00
13) Acenaphthene-d10	16.42	164	155474	5.00	ng	0.00
17) Phenanthrene-d10	19.77	188	296849	5.00	ng	0.00
20) Chrysene-d12	23.66	240	344741	5.00	ng	0.00
27) Perylene-d12	25.34	264	320485	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.50	112	124202	4.56	ng	0.00
4) Phenol-d6	8.62	99	164422	4.87	ng	0.00
8) Nitrobenzene-d5	10.61	82	78538	3.20	ng	0.00
14) 2,4,6-Tribromophenol	18.32	330	19905	6.44	ng	0.00
15) 2-Fluorobiphenyl	14.88	172	212437	4.38	ng	0.00
22) Terphenyl-d14	22.30	244	247006	7.51	ng	0.00
Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.32	42	72455	5.25	ng	# 90
5) Phenol	8.65	94	175818	5.17	ng	# 22
6) bis(2-Chloroethyl)ether	8.83	93	132628	5.19	ng	# 100
9) Nitrobenzene	10.65	77	99030	3.58	ng	# 100
10) 2,4-Dimethylphenol	11.59	122	102772	4.46	ng	97
11) 2,4-Dichlorophenol	11.97	162	80268	4.88	ng	94
12) Hexachlorobutadiene	12.64	225	50139	4.68	ng	# 100
16) 2,4-Dinitrophenol	16.65	184	3739	3.52	ng	96
18) Hexachlorobenzene	19.00	284	62511	4.25	ng	89
19) Pentachlorophenol	19.44	266	13980	5.09	ng	98
21) Benzidine	21.96	184	201863	7.10	ng	99
23) Benzo(a)anthracene	23.64	228	1848756	7.59	ng	97
24) 3,3'-Dichlorobenzidine	23.63	252	152449	6.53	ng	99
25) Chrysene	23.69	228	1665875	4.12	ng	95
26) Indeno(1,2,3-cd)pyrene	26.75	276	490874	2.58	ng	# 100
28) Benzo(b)fluoranthene	24.92	252	409266	14.56	ng	98
29) Benzo(k)fluoranthene	24.94	252	468995	7.04	ng	97
30) Benzo(a)pyrene	25.27	252	385236	10.65	ng	99
31) Dibenzo(a,h)anthracene	26.78	278	393895	12.53	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021615\
Data File : BE089570.D
Acq On : 16 Feb 2015 15:03
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC005

Quant Time: Feb 16 17:17:48 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIM-BE021615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 16 17:05:37 2015
Response via : Initial Calibration

