

Data Path : Z:\HPCHEM1\BNA E\Data\BE021615\
 Data File : BE089577.D
 Acq On : 16 Feb 2015 20:25
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021615

Quant Time: Feb 17 09:56:55 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIM-BE021615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 17 01:13:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	n-Nitrosodimethylamine	0.743	0.660	11.2	87	0.00
3 S	2-Fluorophenol	1.415	1.364	3.6	95	0.00
4 S	Phenol-d6	1.894	1.803	4.8	96	0.00
5 C	Phenol	2.032	1.968	3.1	96	0.00
6	bis(2-Chloroethyl)ether	1.550	1.530	1.3	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.206	0.200	2.9	101	0.00
9	Nitrobenzene	0.242	0.242	0.0	103	0.00
10	2,4-Dimethylphenol	0.287	0.278	3.1	96	0.00
11 C	2,4-Dichlorophenol	0.225	0.220	2.2	97	0.00
12	Hexachlorobutadiene	0.153	0.153	0.0	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.107	0.104	2.8	106	0.00
15 S	2-Fluorobiphenyl	1.348	1.351	-0.2	105	0.00
16 P	2,4-Dinitrophenol	0.017	0.016#	5.9	131	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
18	Hexachlorobenzene	0.211	0.215	-1.9	112	0.00
19 C	Pentachlorophenol	0.036	0.033	8.3	111	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
21	Benzidine	0.396	0.264	33.3#	66	0.00
22 S	Terphenyl-d14	0.679	0.700	-3.1	105	0.00
23	Benzo(a)anthracene	4.782	4.559	4.7	96	0.00
24	3,3'-Dichlorobenzidine	0.317	0.266	16.1	82	0.00
25	Chrysene	4.805	4.858	-1.1	104	0.00
26	Indeno(1,2,3-cd)pyrene	1.101	0.937	14.9	80	0.00
27 I	Perylene-d12	1.000	1.000	0.0	93	0.00
28	Benzo(b)fluoranthene	0.971	0.880	9.4	83	0.00
29	Benzo(k)fluoranthene	1.339	1.422	-6.2	95	0.00
30 C	Benzo(a)pyrene	0.928	0.846	8.8	81	0.00
31	Dibenzo(a,h)anthracene	0.930	0.826	11.2	77	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0