

Data Path : Z:\HPCHEM1\BNA E\Data\BE021615\
 Data File : BE089577.D
 Acq On : 16 Feb 2015 20:25
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021615

Quant Time: Feb 17 09:56:55 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIM-BE021615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 17 01:13:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	n-Nitrosodimethylamine	1.000	0.881	11.9	87	0.00
3 S	2-Fluorophenol	1.000	0.964	3.6	95	0.00
4 S	Phenol-d6	1.000	0.952	4.8	96	0.00
5 C	Phenol	1.000	0.968	3.2	96	0.00
6	bis(2-Chloroethyl)ether	1.000	0.987	1.3	99	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
8 S	Nitrobenzene-d5	1.000	0.974	2.6	101	0.00
9	Nitrobenzene	1.000	0.956	4.4	103	0.00
10	2,4-Dimethylphenol	1.000	0.970	3.0	96	0.00
11 C	2,4-Dichlorophenol	1.000	0.979	2.1	97	0.00
12	Hexachlorobutadiene	1.000	1.002	-0.2	101	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	1.000	0.936	6.4	106	0.00
15 S	2-Fluorobiphenyl	1.000	1.002	-0.2	105	0.00
16 P	2,4-Dinitrophenol	1.000	1.137	-13.7	131	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	112	0.00
18	Hexachlorobenzene	1.000	1.020	-2.0	112	0.00
19 C	Pentachlorophenol	1.000	0.913	8.7	111	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
21	Benzidine	1.000	0.685	31.5#	66	0.00
22 S	Terphenyl-d14	1.000	1.031	-3.1	105	0.00
23	Benzo(a)anthracene	1.000	0.953	4.7	96	0.00
24	3,3'-Dichlorobenzidine	1.000	0.800	20.0	82	0.00
25	Chrysene	1.000	1.011	-1.1	104	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.795	20.5	80	0.00
27 I	Perylene-d12	5.000	5.000	0.0	93	0.00
28	Benzo(b)fluoranthene	1.000	0.843	15.7	83	0.00
29	Benzo(k)fluoranthene	1.000	1.062	-6.2	95	0.00
30 C	Benzo(a)pyrene	1.000	0.849	15.1	81	0.00
31	Dibenzo(a,h)anthracene	1.000	0.820	18.0	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0