

Data Path : Z:\HPCHEM1\BNA E\Data\BE021615\
 Data File : BE089582.D
 Acq On : 16 Feb 2015 23:42
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Feb 17 02:05:52 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIM-BE021615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 17 01:13:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	83	0.00
2	n-Nitrosodimethylamine	1.000	0.828	17.2	66	0.00
3 S	2-Fluorophenol	1.000	0.999	0.1	81	0.00
4 S	Phenol-d6	1.000	0.977	2.3	80	0.00
5 C	Phenol	1.000	0.979	2.1	79	0.00
6	bis(2-Chloroethyl)ether	1.000	0.961	3.9	79	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	83	0.00
8 S	Nitrobenzene-d5	1.000	1.195	-19.5	101	0.00
9	Nitrobenzene	1.000	1.099	-9.9	100	0.00
10	2,4-Dimethylphenol	1.000	0.984	1.6	80	0.00
11 C	2,4-Dichlorophenol	1.000	1.032	-3.2	84	0.00
12	Hexachlorobutadiene	1.000	1.005	-0.5	83	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	1.000	1.051	-5.1	97	0.00
15 S	2-Fluorobiphenyl	1.000	1.019	-1.9	85	0.00
16 P	2,4-Dinitrophenol	1.000	1.334	-33.4#	144	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	88	0.00
18	Hexachlorobenzene	1.000	1.028	-2.8	89	0.00
19 C	Pentachlorophenol	1.000	1.128	-12.8	118	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
21	Benzidine	1.000	0.978	2.2	98	0.00
22 S	Terphenyl-d14	1.000	1.017	-1.7	91	0.00
23	Benzo(a)anthracene	1.000	1.013	-1.3	89	0.00
24	3,3'-Dichlorobenzidine	1.000	0.932	6.8	89	0.00
25	Chrysene	1.000	1.012	-1.2	91	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.970	3.0	90	0.00
27 I	Perylene-d12	5.000	5.000	0.0	91	0.00
28	Benzo(b)fluoranthene	1.000	0.911	8.9	89	0.00
29	Benzo(k)fluoranthene	1.000	1.045	-4.5	91	0.00
30 C	Benzo(a)pyrene	1.000	0.924	7.6	88	0.00
31	Dibenzo(a,h)anthracene	1.000	0.940	6.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0