

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089597.D
 Acq On : 17 Feb 2015 16:28
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Feb 18 01:35:33 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 00:56:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.33	152	54763	5.00	ng	0.00
7) Naphthalene-d8	12.24	136	220033	5.00	ng	0.00
13) Acenaphthene-d10	16.41	164	107332	5.00	ng	0.00
17) Phenanthrene-d10	19.77	188	208684	5.00	ng	0.00
20) Chrysene-d12	23.65	240	216299	5.00	ng	0.00
27) Perylene-d12	25.34	264	188362	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.51	112	11607	0.76	ng	0.00
4) Phenol-d6	8.62	99	15249	0.74	ng	0.00
8) Nitrobenzene-d5	10.60	82	7088	0.78	ng	0.00
14) 2,4,6-Tribromophenol	18.31	330	1362	0.61	ng	0.00
15) 2-Fluorobiphenyl	14.88	172	23653	0.82	ng	0.00
22) Terphenyl-d14	22.30	244	24341	0.83	ng	0.00

Target Compounds				Qvalue		
2) n-Nitrosodimethylamine	3.37	42	5666m	0.72	ng	
5) Phenol	8.65	94	16622	0.76	ng	97
6) bis(2-Chloroethyl)ether	8.82	93	13468	0.80	ng	# 100
9) Nitrobenzene	10.65	77	8714	0.81	ng	99
10) 2,4-Dimethylphenol	11.58	122	8923	0.72	ng	98
11) 2,4-Dichlorophenol	11.96	162	7108	0.73	ng	98
12) Hexachlorobutadiene	12.63	225	5584	0.83	ng	99
16) 2,4-Dinitrophenol	16.66	184	251	0.66	ng	93
18) Hexachlorobenzene	18.99	284	7398	0.84	ng	99
19) Pentachlorophenol	19.44	266	872	0.58	ng	99
21) Benzidine	21.96	184	7877	0.49	ng	99
23) Benzo(a)anthracene	23.64	228	147286	0.73	ng	100
24) 3,3'-Dichlorobenzidine	23.64	252	6561	0.50	ng	99
25) Chrysene	23.69	228	174282	0.83	ng	99
26) Indeno(1,2,3-cd)pyrene	26.75	276	27775	0.61	ng	# 99
28) Benzo(b)fluoranthene	24.91	252	19569	0.55	ng	98
29) Benzo(k)fluoranthene	24.94	252	45508	0.90	ng	99
30) Benzo(a)pyrene	25.28	252	20834	0.62	ng	99
31) Dibenzo(a,h)anthracene	26.78	278	21324	0.63	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
Data File : BE089597.D
Acq On : 17 Feb 2015 16:28
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Feb 18 01:35:33 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 00:56:39 2015
Response via : Initial Calibration

