

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089599.D
 Acq On : 17 Feb 2015 17:45
 Operator : TP/IZ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC002

Quant Time: Feb 18 01:11:27 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 00:56:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	60840	5.00	ng	0.00
7) Naphthalene-d8	12.24	136	244208	5.00	ng	0.00
13) Acenaphthene-d10	16.42	164	121904	5.00	ng	0.00
17) Phenanthrene-d10	19.77	188	237165	5.00	ng	0.00
20) Chrysene-d12	23.66	240	251825	5.00	ng	0.00
27) Perylene-d12	25.34	264	215451	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.51	112	32064	1.88	ng	0.00
4) Phenol-d6	8.62	99	42546	1.87	ng	0.00
8) Nitrobenzene-d5	10.61	82	21898	2.18	ng	0.00
14) 2,4,6-Tribromophenol	18.31	330	4696	1.86	ng	0.00
15) 2-Fluorobiphenyl	14.88	172	65164	1.99	ng	0.00
22) Terphenyl-d14	22.30	244	70383	2.06	ng	0.00

Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.33	42	14750	1.69	ng	# 54
5) Phenol	8.65	94	46232	1.89	ng	97
6) bis(2-Chloroethyl)ether	8.83	93	35972	1.92	ng	# 99
9) Nitrobenzene	10.65	77	28108	2.35	ng	99
10) 2,4-Dimethylphenol	11.58	122	26598	1.93	ng	98
11) 2,4-Dichlorophenol	11.96	162	21486	1.98	ng	99
12) Hexachlorobutadiene	12.63	225	14914	1.99	ng	99
16) 2,4-Dinitrophenol	16.65	184	940	2.18	ng	87
18) Hexachlorobenzene	19.00	284	20533	2.05	ng	# 83
19) Pentachlorophenol	19.44	266	3336	1.97	ng	98
21) Benzidine	21.96	184	29602	1.57	ng	99
23) Benzo(a)anthracene	23.64	228	457590	1.94	ng	98
24) 3,3'-Dichlorobenzidine	23.64	252	26739	1.76	ng	# 99
25) Chrysene	23.69	228	480851	1.98	ng	98
26) Indeno(1,2,3-cd)pyrene	26.75	276	93945	1.76	ng	# 100
28) Benzo(b)fluoranthene	24.92	252	78321	1.94	ng	99
29) Benzo(k)fluoranthene	24.94	252	129007	2.23	ng	100
30) Benzo(a)pyrene	25.28	252	75743	1.96	ng	99
31) Dibenzo(a,h)anthracene	26.78	278	72886	1.89	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
Data File : BE089599.D
Acq On : 17 Feb 2015 17:45
Operator : TP/IZ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC002

Quant Time: Feb 18 01:11:27 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 00:56:39 2015
Response via : Initial Calibration

