

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089607.D
 Acq On : 17 Feb 2015 23:37
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Feb 18 14:44:48 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 14:11:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	52276	5.00	ng	0.00
7) Naphthalene-d8	12.24	136	213056	5.00	ng	0.00
13) Acenaphthene-d10	16.42	164	107290	5.00	ng	0.00
17) Phenanthrene-d10	19.77	188	213509	5.00	ng	0.00
20) Chrysene-d12	23.65	240	218550	5.00	ng	0.00
27) Perylene-d12	25.34	264	191900	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.50	112	11883	0.87	ng	0.00
4) Phenol-d6	8.62	99	15485	0.86	ng	0.00
8) Nitrobenzene-d5	10.60	82	8490	0.91	ng	0.00
14) 2,4,6-Tribromophenol	18.31	330	1677	0.86	ng	0.00
15) 2-Fluorobiphenyl	14.88	172	23379	0.81	ng	0.00
22) Terphenyl-d14	22.30	244	26066	0.88	ng	0.00
Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.36	42	4538	0.70	ng	# 68
5) Phenol	8.65	94	16536	0.86	ng	88
6) bis(2-Chloroethyl)ether	8.83	93	12631	0.82	ng	# 99
9) Nitrobenzene	10.65	77	10080	0.87	ng	98
10) 2,4-Dimethylphenol	11.58	122	9511	0.87	ng	100
11) 2,4-Dichlorophenol	11.96	162	7784	0.88	ng	99
12) Hexachlorobutadiene	12.63	225	5376	0.82	ng	100
16) 2,4-Dinitrophenol	16.66	184	278	0.89	ng	97
18) Hexachlorobenzene	18.99	284	7466	0.81	ng	95
19) Pentachlorophenol	19.44	266	1103	0.91	ng	99
21) Benzidine	21.96	184	12204	0.99	ng	100
23) Benzo(a)anthracene	23.64	228	167711	0.90	ng	98
24) 3,3'-Dichlorobenzidine	23.63	252	10238	1.07	ng	99
25) Chrysene	23.69	228	172731	0.81	ng	99
26) Indeno(1,2,3-cd)pyrene	26.75	276	35300	0.90	ng	# 99
28) Benzo(b)fluoranthene	24.91	252	29169	0.90	ng	99
29) Benzo(k)fluoranthene	24.94	252	44884	0.81	ng	99
30) Benzo(a)pyrene	25.28	252	27761	0.88	ng	98
31) Dibenzo(a,h)anthracene	26.78	278	27692	0.89	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
Data File : BE089607.D
Acq On : 17 Feb 2015 23:37
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Quant Time: Feb 18 14:44:48 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 14:11:34 2015
Response via : Initial Calibration

