

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089602.D
 Acq On : 17 Feb 2015 19:45
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021715

Quant Time: Feb 18 14:32:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 14:11:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	n-Nitrosodimethylamine	0.587	0.556	5.3	99	0.00
3 S	2-Fluorophenol	1.301	1.293	0.6	111	0.00
4 S	Phenol-d6	1.714	1.678	2.1	110	0.00
5 C	Phenol	1.846	1.811	1.9	109	0.00
6	bis(2-Chloroethyl)ether	1.472	1.394	5.3	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.217	0.222	-2.3	123	0.00
9	Nitrobenzene	0.260	0.273	-5.0	122	0.00
10	2,4-Dimethylphenol	0.257	0.258	-0.4	114	0.00
11 C	2,4-Dichlorophenol	0.207	0.207	0.0	114	0.00
12	Hexachlorobutadiene	0.155	0.149	3.9	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.094	0.084	10.6	122	0.00
15 S	2-Fluorobiphenyl	1.337	1.292	3.4	112	0.00
16 P	2,4-Dinitrophenol	0.018	0.016#	11.1	131	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
18	Hexachlorobenzene	0.216	0.206	4.6	116	0.00
19 C	Pentachlorophenol	0.033	0.027	18.2	113	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
21	Benzidine	0.281	0.219	22.1	105	0.00
22 S	Terphenyl-d14	0.675	0.666	1.3	116	0.00
23	Benzo(a)anthracene	4.262	4.184	1.8	115	0.00
24	3,3'-Dichlorobenzidine	0.206	0.218	-5.8	123	0.00
25	Chrysene	4.880	4.633	5.1	107	0.00
26	Indeno(1,2,3-cd)pyrene	0.756	0.805	-6.5	110	0.00
27 I	Perylene-d12	1.000	1.000	0.0	112	0.00
28	Benzo(b)fluoranthene	0.699	0.734	-5.0	113	0.00
29	Benzo(k)fluoranthene	1.346	1.360	-1.0	107	0.00
30 C	Benzo(a)pyrene	0.688	0.723	-5.1	113	0.00
31	Dibenzo(a,h)anthracene	0.660	0.714	-8.2	111	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0