

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089602.D
 Acq On : 17 Feb 2015 19:45
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021715

Quant Time: Feb 18 14:32:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 14:11:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	113	0.00
2	n-Nitrosodimethylamine	1.000	0.880	12.0	99	0.00
3 S	2-Fluorophenol	1.000	0.994	0.6	111	0.00
4 S	Phenol-d6	1.000	0.979	2.1	110	0.00
5 C	Phenol	1.000	0.981	1.9	109	0.00
6	bis(2-Chloroethyl)ether	1.000	0.947	5.3	106	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	115	0.00
8 S	Nitrobenzene-d5	1.000	0.984	1.6	123	0.00
9	Nitrobenzene	1.000	0.974	2.6	122	0.00
10	2,4-Dimethylphenol	1.000	1.003	-0.3	114	0.00
11 C	2,4-Dichlorophenol	1.000	1.001	-0.1	114	0.00
12	Hexachlorobutadiene	1.000	0.963	3.7	110	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	1.000	0.899	10.1	122	0.00
15 S	2-Fluorobiphenyl	1.000	0.966	3.4	112	0.00
16 P	2,4-Dinitrophenol	1.000	0.999	0.1	131	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	122	0.00
18	Hexachlorobenzene	1.000	0.958	4.2	116	0.00
19 C	Pentachlorophenol	1.000	0.931	6.9	113	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	115	0.00
21	Benzidine	1.000	0.831	16.9	105	0.00
22 S	Terphenyl-d14	1.000	0.987	1.3	116	0.00
23	Benzo(a)anthracene	1.000	0.982	1.8	115	0.00
24	3,3'-Dichlorobenzidine	1.000	1.008	-0.8	123	0.00
25	Chrysene	1.000	0.949	5.1	107	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.894	10.6	110	0.00
27 I	Perylene-d12	5.000	5.000	0.0	112	0.00
28	Benzo(b)fluoranthene	1.000	0.880	12.0	113	0.00
29	Benzo(k)fluoranthene	1.000	0.931	6.9	107	0.00
30 C	Benzo(a)pyrene	1.000	0.884	11.6	113	0.00
31	Dibenzo(a,h)anthracene	1.000	0.888	11.2	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0