

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089607.D
 Acq On : 17 Feb 2015 23:37
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Feb 18 14:44:48 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 14:11:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	n-Nitrosodimethylamine	0.587	0.434	26.1#	63	0.00
3 S	2-Fluorophenol	1.301	1.137	12.6	80	0.00
4 S	Phenol-d6	1.714	1.481	13.6	80	0.00
5 C	Phenol	1.846	1.582	14.3	78	0.00
6	bis(2-Chloroethyl)ether	1.472	1.208	17.9	75	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.217	0.199	8.3	92	0.00
9	Nitrobenzene	0.260	0.237	8.8	88	0.00
10	2,4-Dimethylphenol	0.257	0.223	13.2	83	0.00
11 C	2,4-Dichlorophenol	0.207	0.183	11.6	84	0.00
12	Hexachlorobutadiene	0.155	0.126	18.7	78	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.094	0.078	17.0	96	0.00
15 S	2-Fluorobiphenyl	1.337	1.090	18.5	80	0.00
16 P	2,4-Dinitrophenol	0.018	0.013#	27.8#	89	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
18	Hexachlorobenzene	0.216	0.175	19.0	83	0.00
19 C	Pentachlorophenol	0.033	0.026	21.2#	93	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
21	Benzidine	0.281	0.279	0.7	112	0.00
22 S	Terphenyl-d14	0.675	0.596	11.7	87	0.00
23	Benzo(a)anthracene	4.262	3.837	10.0	89	0.00
24	3,3'-Dichlorobenzidine	0.206	0.234	-13.6	111	0.00
25	Chrysene	4.880	3.952	19.0	77	0.00
26	Indeno(1,2,3-cd)pyrene	0.756	0.808	-6.9	93	0.00
27 I	Perylene-d12	1.000	1.000	0.0	94	0.00
28	Benzo(b)fluoranthene	0.699	0.760	-8.7	99	0.00
29	Benzo(k)fluoranthene	1.346	1.169	13.2	78	0.00
30 C	Benzo(a)pyrene	0.688	0.723	-5.1	95	0.00
31	Dibenzo(a,h)anthracene	0.660	0.722	-9.4	95	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1