

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021715\
 Data File : BE089607.D
 Acq On : 17 Feb 2015 23:37
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Feb 18 14:44:48 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE021715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 14:11:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	93	0.00
2	n-Nitrosodimethylamine	1.000	0.697	30.3#	63	0.00
3 S	2-Fluorophenol	1.000	0.873	12.7	80	0.00
4 S	Phenol-d6	1.000	0.864	13.6	80	0.00
5 C	Phenol	1.000	0.857	14.3	78	0.00
6	bis(2-Chloroethyl)ether	1.000	0.821	17.9	75	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	96	0.00
8 S	Nitrobenzene-d5	1.000	0.905	9.5	92	0.00
9	Nitrobenzene	1.000	0.866	13.4	88	0.00
10	2,4-Dimethylphenol	1.000	0.867	13.3	83	0.00
11 C	2,4-Dichlorophenol	1.000	0.882	11.8	84	0.00
12	Hexachlorobutadiene	1.000	0.816	18.4	78	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	1.000	0.860	14.0	96	0.00
15 S	2-Fluorobiphenyl	1.000	0.815	18.5	80	0.00
16 P	2,4-Dinitrophenol	1.000	0.895	10.5	89	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	104	0.00
18	Hexachlorobenzene	1.000	0.811	18.9	83	0.00
19 C	Pentachlorophenol	1.000	0.909	9.1	93	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	97	0.00
21	Benzidine	1.000	0.992	0.8	112	0.00
22 S	Terphenyl-d14	1.000	0.884	11.6	87	0.00
23	Benzo(a)anthracene	1.000	0.900	10.0	89	0.00
24	3,3'-Dichlorobenzidine	1.000	1.070	-7.0	111	0.00
25	Chrysene	1.000	0.810	19.0	77	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.896	10.4	93	0.00
27 I	Perylene-d12	5.000	5.000	0.0	94	0.00
28	Benzo(b)fluoranthene	1.000	0.902	9.8	99	0.00
29	Benzo(k)fluoranthene	1.000	0.811	18.9	78	0.00
30 C	Benzo(a)pyrene	1.000	0.884	11.6	95	0.00
31	Dibenzo(a,h)anthracene	1.000	0.895	10.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0