

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021716\
 Data File : BE091411.D
 Acq On : 17 Feb 2016 19:20
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 18 00:00:41 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	-0.01
2 S	2-Fluorophenol	0.896	1.113	-24.2	138	-0.04
3 S	Phenol-d6	1.352	1.724	-27.5#	131	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	102	-0.01
5 S	Nitrobenzene-d5	0.241	0.332	-37.8#	162#	-0.03
6	Naphthalene	1.128	1.057	6.3	108	-0.02
7	2-Methylnaphthalene	0.609	0.655	-7.6	124	-0.02
8 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
9 S	2,4,6-Tribromophenol	0.153	0.216	-41.2#	180#	-0.01
10 S	2-Fluorobiphenyl	1.301	1.260	3.2	110	-0.02
11	Acenaphthylene	7.133	8.271	-16.0	137	0.00
12	Acenaphthene	1.373	1.265	7.9	108	0.00
13	Fluorene	1.772	1.709	3.6	109	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	96	-0.01
15	Phenanthrene	1.313	1.216	7.4	101	0.00
16	Anthracene	1.100	1.201	-9.2	116	-0.01
17	Fluoranthene	1.203	1.373	-14.1	126	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
19	Pyrene	1.083	1.185	-9.4	124	0.00
20 S	Terphenyl-d14	0.740	0.806	-8.9	120	0.00
21	Benzo(a)anthracene	0.926	1.180	-27.4#	152#	0.00
22	Chrysene	1.371	1.218	11.2	102	0.00
23	Indeno(1,2,3-cd)pyrene	0.929	1.256	-35.2#	148	-0.01
24 I	Perylene-d12	1.000	1.000	0.0	111	0.00
25	Benzo(b)fluoranthene	1.048	1.123	-7.2	129	0.00
26	Benzo(k)fluoranthene	1.187	1.269	-6.9	133	0.00
27 C	Benzo(a)pyrene	0.939	1.132	-20.6#	148	0.00
28	Dibenzo(a,h)anthracene	0.827	1.168	-41.2#	173#	-0.01
29	Benzo(g,h,i)perylene	1.055	1.222	-15.8	143	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1