

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021815\
 Data File : BE089621.D
 Acq On : 18 Feb 2015 16:56
 Operator : TP/IZ
 Sample : MDL-01-W
 Misc : MDL-WATER-0.1/0.2
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-01-W

Manual Integrations
 APPROVED

apatel
 2/19/2015 1:49:00 PM

Quant Time: Feb 19 14:03:23 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	37507	5.00	ng	0.00
4) Naphthalene-d8	12.25	136	149976	5.00	ng	0.00
8) Acenaphthene-d10	16.42	164	72888	5.00	ng	0.00
13) Phenanthrene-d10	19.77	188	143008	5.00	ng	0.00
19) Chrysene-d12	23.66	240	131712	5.00	ng	0.00
25) Perylene-d12	25.34	264	107711	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	10.61	82	31962	4.90	ng	0.00
9) 2-Fluorobiphenyl	14.88	172	112381	5.76	ng	0.00
21) Terphenyl-d14	22.30	244	224148	11.84	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.75	88	143	0.03	ng	# 83
6) Naphthalene	12.29	128	1307	0.04	ng	# 88
7) 2-Methylnaphthalene	13.95	142	751	0.04	ng	94
10) Acenaphthylene	16.07	152	4666	0.04	ng	99
11) Acenaphthene	16.50	154	938	0.05	ng	93
12) Fluorene	17.74	166	1139	0.05	ng	98
14) Hexachlorobenzene	19.00	284	390	0.06	ng	92
15) Pentachlorophenol	19.44	266	43	0.02	ng	# 73
16) Phenanthrene	19.82	178	2310	0.07	ng	99
17) Anthracene	19.91	178	1547	0.05	ng	98
18) Fluoranthene	21.72	202	1783	0.05	ng	99
20) Pyrene	22.04	202	1932	0.06	ng	99
22) Benzo(a)anthracene	23.65	228	7245	0.06	ng	# 92
23) Chrysene	23.69	228	10913	0.08	ng	96
24) Indeno(1,2,3-cd)pyrene	26.74	276	2594	0.20	ng	# 88
26) Benzo(b)fluoranthene	24.91	252	581	0.29	ng	# 79
27) Benzo(k)fluoranthene	24.94	252	1257m	0.04	ng	
28) Benzo(a)pyrene	25.27	252	568	0.26	ng	# 74
29) Dibenzo(a,h)anthracene	26.77	278	385	0.25	ng	# 57
30) Benzo(g,h,i)perylene	27.17	276	2996	0.16	ng	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021815\
Data File : BE089621.D
Acq On : 18 Feb 2015 16:56
Operator : TP/IZ
Sample : MDL-01-W
Misc : MDL-WATER-0.1/0.2
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-01-W

Manual Integrations
APPROVED

apatel
2/19/2015 1:49:00 PM

Quant Time: Feb 19 14:03:23 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 13:01:39 2015
Response via : Initial Calibration

