

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021815\
 Data File : BE089626.D
 Acq On : 18 Feb 2015 20:12
 Operator : TP/IZ
 Sample : MDL-06-W
 Misc : MDL-WATER-0.1/0.2
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-06-W

Manual Integrations
 APPROVED

apatel
 2/19/2015 1:49:29 PM

Quant Time: Feb 19 05:02:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	37157	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	147268	5.00	ng	0.00
8) Acenaphthene-d10	16.42	164	72762	5.00	ng	0.00
13) Phenanthrene-d10	19.77	188	143226	5.00	ng	0.00
19) Chrysene-d12	23.66	240	127774	5.00	ng	0.00
25) Perylene-d12	25.34	264	100613	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	10.61	82	65970	10.31	ng	0.00
9) 2-Fluorobiphenyl	14.88	172	176896	9.08	ng	0.00
21) Terphenyl-d14	22.30	244	209849	11.42	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.74	88	326	0.07	ng	# 80
6) Naphthalene	12.29	128	2586	0.08	ng	# 93
7) 2-Methylnaphthalene	13.95	142	1406	0.07	ng	96
10) Acenaphthylene	16.07	152	7578	0.06	ng	100
11) Acenaphthene	16.50	154	1432	0.08	ng	89
12) Fluorene	17.74	166	1576	0.07	ng	98
14) Hexachlorobenzene	19.00	284	501	0.08	ng	# 92
15) Pentachlorophenol	19.44	266	31	Below Cal		# 67
16) Phenanthrene	19.82	178	2756	0.08	ng	99
17) Anthracene	19.91	178	1937	0.06	ng	99
18) Fluoranthene	21.72	202	1967	0.06	ng	99
20) Pyrene	22.04	202	2087	0.07	ng	99
22) Benzo(a)anthracene	23.65	228	7092m	0.06	ng	
23) Chrysene	23.69	228	11496	0.09	ng	95
24) Indeno(1,2,3-cd)pyrene	26.75	276	2155	0.19	ng	95
26) Benzo(b)fluoranthene	24.91	252	545	0.29	ng	# 78
27) Benzo(k)fluoranthene	24.94	252	1067m	0.04	ng	
28) Benzo(a)pyrene	25.27	252	543	0.26	ng	# 73
29) Dibenzo(a,h)anthracene	26.77	278	380	0.25	ng	# 52
30) Benzo(g,h,i)perylene	27.16	276	2440	0.16	ng	# 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE021815\
Data File : BE089626.D
Acq On : 18 Feb 2015 20:12
Operator : TP/IZ
Sample : MDL-06-W
Misc : MDL-WATER-0.1/0.2
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-06-W

Manual Integrations
APPROVED

apatel
2/19/2015 1:49:29 PM

Quant Time: Feb 19 05:02:55 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 13:01:39 2015
Response via : Initial Calibration

