

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021815\
 Data File : BE089643.D
 Acq On : 20 Feb 2015 1:48
 Operator : TP/IZ
 Sample : MDL-05-W
 Misc : MDL-WATER-0.2/0.4
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MDL-05-W

Quant Time: Feb 20 02:34:57 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	30569	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	121870	5.00	ng	0.00
8) Acenaphthene-d10	16.42	164	61361	5.00	ng	0.00
13) Phenanthrene-d10	19.77	188	119545	5.00	ng	0.00
19) Chrysene-d12	23.66	240	109826	5.00	ng	0.00
25) Perylene-d12	25.34	264	87233	5.00	ng	0.00

System Monitoring Compounds						
5) Nitrobenzene-d5	10.61	82	57995	10.95	ng	0.00
9) 2-Fluorobiphenyl	14.88	172	166599	10.14	ng	0.00
21) Terphenyl-d14	22.30	244	201629	12.77	ng	0.00

Target Compounds				Ovalue		
2) 1,4-Dioxane	2.72	88	594	0.16	ng	# 83
3) n-Nitrosodimethylamine	3.48	42	430	0.12	ng	# 95
6) Naphthalene	12.29	128	4752	0.17	ng	97
7) 2-Methylnaphthalene	13.95	142	2633	0.16	ng	98
10) Acenaphthylene	16.07	152	14642	0.15	ng	100
11) Acenaphthene	16.49	154	2646	0.16	ng	96
12) Fluorene	17.74	166	2981	0.16	ng	98
14) Hexachlorobenzene	19.00	284	934	0.18	ng	97
15) Pentachlorophenol	19.44	266	55	0.05	ng	# 78
16) Phenanthrene	19.81	178	4998	0.17	ng	98
17) Anthracene	19.91	178	3735	0.15	ng	99
18) Fluoranthene	21.72	202	3961	0.14	ng	99
20) Pyrene	22.04	202	4290	0.16	ng	99
22) Benzo(a)anthracene	23.65	228	12269	0.13	ng	96
23) Chrysene	23.69	228	21714	0.20	ng	98
24) Indeno(1,2,3-cd)pyrene	26.74	276	4123	0.22	ng	95
26) Benzo(b)fluoranthene	24.91	252	944	0.31	ng	# 88
27) Benzo(k)fluoranthene	24.94	252	2807m	0.11	ng	
28) Benzo(a)pyrene	25.27	252	1018	0.29	ng	# 84
29) Dibenzo(a,h)anthracene	26.77	278	622	0.26	ng	# 73
30) Benzo(g,h,i)perylene	27.16	276	5971	0.21	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021815\
Data File : BE089643.D
Acq On : 20 Feb 2015 1:48
Operator : TP/IZ
Sample : MDL-05-W
Misc : MDL-WATER-0.2/0.4
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MDL-05-W

Quant Time: Feb 20 02:34:57 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE021815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 18 13:01:39 2015
Response via : Initial Calibration

