

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021815\
 Data File : BE089617.D
 Acq On : 18 Feb 2015 6:05
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE021815

Quant Time: Feb 18 13:10:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	0.970	3.0	99	0.00
3	n-Nitrosodimethylamine	1.000	0.998	0.2	99	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	103	0.00
5 S	Nitrobenzene-d5	1.000	0.961	3.9	102	0.00
6	Naphthalene	1.000	0.957	4.3	97	0.00
7	2-Methylnaphthalene	1.000	0.959	4.1	98	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.00
9 S	2-Fluorobiphenyl	1.000	0.965	3.5	98	0.00
10	Acenaphthylene	1.000	0.974	2.6	99	0.00
11	Acenaphthene	1.000	0.955	4.5	98	0.00
12	Fluorene	1.000	0.955	4.5	97	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	104	0.00
14	Hexachlorobenzene	1.000	0.957	4.3	99	0.00
15	Pentachlorophenol	1.000	0.914	8.6	94	0.00
16	Phenanthrene	1.000	0.954	4.6	99	0.00
17	Anthracene	1.000	0.967	3.3	100	0.00
18	Fluoranthene	1.000	0.959	4.1	100	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
20	Pyrene	1.000	0.964	3.6	100	0.00
21 S	Terphenyl-d14	1.000	0.950	5.0	101	0.00
22	Benzo(a)anthracene	1.000	0.941	5.9	102	0.00
23	Chrysene	1.000	0.960	4.0	102	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.873	12.7	101	0.00
25 I	Perylene-d12	5.000	5.000	0.0	106	0.00
26	Benzo(b)fluoranthene	1.000	0.844	15.6	99	0.00
27	Benzo(k)fluoranthene	1.000	1.018	-1.8	102	0.00
28 C	Benzo(a)pyrene	1.000	0.870	13.0	101	0.00
29	Dibenzo(a,h)anthracene	1.000	0.874	12.6	101	0.00
30	Benzo(g,h,i)perylene	1.000	0.898	10.2	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0