

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021815\
 Data File : BE089632.D
 Acq On : 19 Feb 2015 00:53
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 19 05:13:58 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	1,4-Dioxane	1.000	1.010	-1.0	95	0.00
3	n-Nitrosodimethylamine	1.000	0.999	0.1	92	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
5 S	Nitrobenzene-d5	1.000	0.939	6.1	91	0.00
6	Naphthalene	1.000	1.004	-0.4	93	0.00
7	2-Methylnaphthalene	1.000	1.005	-0.5	94	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	96	0.00
9 S	2-Fluorobiphenyl	1.000	1.008	-0.8	95	0.00
10	Acenaphthylene	1.000	1.002	-0.2	94	0.00
11	Acenaphthene	1.000	0.994	0.6	94	0.00
12	Fluorene	1.000	1.001	-0.1	94	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
14	Hexachlorobenzene	1.000	1.030	-3.0	98	0.00
15	Pentachlorophenol	1.000	0.784	21.6	72	0.00
16	Phenanthrene	1.000	0.999	0.1	95	0.00
17	Anthracene	1.000	1.013	-1.3	96	0.00
18	Fluoranthene	1.000	1.013	-1.3	97	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	97	0.00
20	Pyrene	1.000	1.022	-2.2	96	0.00
21 S	Terphenyl-d14	1.000	0.997	0.3	96	0.00
22	Benzo(a)anthracene	1.000	0.905	9.5	89	0.00
23	Chrysene	1.000	1.043	-4.3	100	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.734	26.6#	73	0.00
25 I	Perylene-d12	5.000	5.000	0.0	95	0.00
26	Benzo(b)fluoranthene	1.000	0.699	30.1#	67	0.00
27	Benzo(k)fluoranthene	1.000	1.063	-6.3	96	0.00
28 C	Benzo(a)pyrene	1.000	0.763	23.7#	75	0.00
29	Dibenzo(a,h)anthracene	1.000	0.724	27.6#	70	0.00
30	Benzo(g,h,i)perylene	1.000	0.848	15.2	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1