

Data Path : Z:\HPCHEM1\BNA E\Data\BE021815\
 Data File : BE089646.D
 Acq On : 20 Feb 2015 4:23
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 20 07:54:15 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.609	0.598	1.8	100	0.00
3	n-Nitrosodimethylamine	0.584	0.564	3.4	96	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
5 S	Nitrobenzene-d5	0.217	0.216	0.5	103	0.00
6	Naphthalene	1.122	1.133	-1.0	101	0.00
7	2-Methylnaphthalene	0.688	0.697	-1.3	101	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
9 S	2-Fluorobiphenyl	1.339	1.362	-1.7	103	0.00
10	Acenaphthylene	8.112	8.251	-1.7	104	0.00
11	Acenaphthene	1.307	1.307	0.0	102	-0.01
12	Fluorene	1.509	1.524	-1.0	103	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
14	Hexachlorobenzene	0.215	0.224	-4.2	108	0.00
15	Pentachlorophenol	0.034	0.024	29.4#	84	0.00
16	Phenanthrene	1.226	1.209	1.4	102	-0.01
17	Anthracene	1.046	1.065	-1.8	105	0.00
18	Fluoranthene	1.169	1.166	0.3	104	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
20	Pvrene	1.233	1.258	-2.0	104	0.00
21 S	Terphenyl-d14	0.719	0.731	-1.7	106	0.00
22	Benzo(a)anthracene	4.411	3.993	9.5	97	0.00
23	Chrysene	4.915	5.085	-3.5	108	0.00
24	Indeno(1,2,3-cd)pyrene	3.394	2.416	28.8#	75	0.00
25 I	Perylene-d12	1.000	1.000	0.0	103	0.00
26	Benzo(b)fluoranthene	0.813	0.503	38.1#	68	0.00
27	Benzo(k)fluoranthene	1.423	1.506	-5.8	103	0.00
28 C	Benzo(a)pyrene	0.802	0.585	27.1#	78	0.00
29	Dibenzo(a,h)anthracene	0.780	0.524	32.8#	71	0.00
30	Benzo(g,h,i)perylene	3.682	3.156	14.3	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1