

Data Path : Z:\HPCHEM1\BNA E\Data\BE021815\
 Data File : BE089646.D
 Acq On : 20 Feb 2015 4:23
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 20 07:54:15 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	0.982	1.8	100	0.00
3	n-Nitrosodimethylamine	1.000	0.965	3.5	96	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	101	0.00
5 S	Nitrobenzene-d5	1.000	0.994	0.6	103	0.00
6	Naphthalene	1.000	1.010	-1.0	101	0.00
7	2-Methylnaphthalene	1.000	1.013	-1.3	101	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.00
9 S	2-Fluorobiphenyl	1.000	1.017	-1.7	103	0.00
10	Acenaphthylene	1.000	1.017	-1.7	104	0.00
11	Acenaphthene	1.000	1.000	0.0	102	-0.01
12	Fluorene	1.000	1.010	-1.0	103	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	104	0.00
14	Hexachlorobenzene	1.000	1.045	-4.5	108	0.00
15	Pentachlorophenol	1.000	0.823	17.7	84	0.00
16	Phenanthrene	1.000	0.986	1.4	102	-0.01
17	Anthracene	1.000	1.018	-1.8	105	0.00
18	Fluoranthene	1.000	0.997	0.3	104	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	105	0.00
20	Pvrene	1.000	1.020	-2.0	104	0.00
21 S	Terphenyl-d14	1.000	1.016	-1.6	106	0.00
22	Benzo(a)anthracene	1.000	0.905	9.5	97	0.00
23	Chrysene	1.000	1.035	-3.5	108	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.703	29.7#	75	0.00
25 I	Perylene-d12	5.000	5.000	0.0	103	0.00
26	Benzo(b)fluoranthene	1.000	0.675	32.5#	68	0.00
27	Benzo(k)fluoranthene	1.000	1.058	-5.8	103	0.00
28 C	Benzo(a)pyrene	1.000	0.742	25.8#	78	0.00
29	Dibenzo(a,h)anthracene	1.000	0.695	30.5#	71	0.00
30	Benzo(g,h,i)perylene	1.000	0.822	17.8	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1