

Data Path : Z:\HPCHEM1\BNA_E\Data\BE021815\
 Data File : Be089619.d
 Acq On : 18 Feb 2015 15:17
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 19 04:34:28 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 18 13:01:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	87	0.00
2	1,4-Dioxane	1.000	0.979	2.1	85	0.00
3	n-Nitrosodimethylamine	1.000	0.965	3.5	81	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	87	0.00
5 S	Nitrobenzene-d5	1.000	0.885	11.5	80	0.00
6	Naphthalene	1.000	1.005	-0.5	87	0.00
7	2-Methylnaphthalene	1.000	0.999	0.1	86	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	89	0.00
9 S	2-Fluorobiphenyl	1.000	1.016	-1.6	88	0.00
10	Acenaphthylene	1.000	1.010	-1.0	88	0.00
11	Acenaphthene	1.000	1.005	-0.5	87	0.00
12	Fluorene	1.000	1.009	-0.9	88	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	89	0.00
14	Hexachlorobenzene	1.000	1.020	-2.0	90	0.00
15	Pentachlorophenol	1.000	0.990	1.0	88	0.00
16	Phenanthrene	1.000	1.018	-1.8	90	0.00
17	Anthracene	1.000	0.998	0.2	88	0.00
18	Fluoranthene	1.000	1.010	-1.0	90	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
20	Pyrene	1.000	1.007	-0.7	91	0.00
21 S	Terphenyl-d14	1.000	0.987	1.3	91	0.00
22	Benzo(a)anthracene	1.000	0.949	5.1	90	0.00
23	Chrysene	1.000	1.014	-1.4	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.834	16.6	83	0.00
25 I	Perylene-d12	5.000	5.000	0.0	94	0.00
26	Benzo(b)fluoranthene	1.000	0.793	20.7	80	0.00
27	Benzo(k)fluoranthene	1.000	1.052	-5.2	93	0.00
28 C	Benzo(a)pyrene	1.000	0.827	17.3	83	0.00
29	Dibenzo(a,h)anthracene	1.000	0.818	18.2	81	0.00
30	Benzo(g,h,i)perylene	1.000	0.895	10.5	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0