


Data Path : Z:\HPCHEM1\BNA_E\Data\BE021815\
 Data File : BE089618.D
 Acq On : 18 Feb 2015 14:38
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE021815.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 18 13:01:39 2015



AutoFind: Scans 3240, 3241, 3242; Background Corrected with Scan 3231

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.7	79484	PASS
68	69	0.00	2	1.9	1637	PASS
69	198	0.00	100	33.3	86379	PASS
70	69	0.00	2	0.6	491	PASS
127	198	10	80	47.1	122018	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	259298	PASS
199	198	5	9	6.3	16369	PASS
275	198	10	60	26.2	67849	PASS
365	198	1	100	3.3	8554	PASS
441	442	0.01	24	15.6	39962	PASS
442	198	50	100	98.9	256567	PASS
443	442	15	24	19.6	50205	PASS