

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021916\
 Data File : BE091413.D
 Acq On : 19 Feb 2016 13:19
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Feb 19 22:37:56 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	43184	5.00	ng	-0.01
4) Naphthalene-d8	9.53	136	213677	5.00	ng	-0.02
8) Acenaphthene-d10	13.35	164	139454	5.00	ng	0.00
14) Phenanthrene-d10	16.08	188	341114	5.00	ng	-0.01
18) Chrysene-d12	20.21	240	385194	5.00	ng	0.00
24) Perylene-d12	22.09	264	393603	5.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	4.92	112	8964	1.07	ng	-0.03
3) Phenol-d6	6.55	99	10923	0.94	ng	-0.02
5) Nitrobenzene-d5	8.15	82	12502	1.11	ng	-0.02
9) 2,4,6-Tribromophenol	14.89	330	4805	1.01	ng	-0.01
10) 2-Fluorobiphenyl	12.00	172	34721	0.96	ng	-0.01
20) Terphenyl-d14	18.61	244	60659	1.06	ng	0.00
Target Compounds						Qvalue
6) Naphthalene	9.58	128	45192	0.94	ng	99
7) 2-Methylnaphthalene	11.11	142	27722	1.03	ng	100
11) Acenaphthylene	13.09	152	216929	1.04	ng	100
12) Acenaphthene	13.41	154	35709	0.93	ng	96
13) Fluorene	14.39	166	48001	0.97	ng	100
15) Phenanthrene	16.12	178	84999	0.95	ng	99
16) Anthracene	16.20	178	78734	1.05	ng	100
17) Fluoranthene	18.09	202	87413	1.03	ng	99
19) Pyrene	18.47	202	91332	1.09	ng	100
21) Benzo(a)anthracene	20.18	228	82481	1.10	ng	98
22) Chrysene	20.24	228	100384	1.00	ng	99
23) Indeno(1,2,3-cd)pyrene	23.69	276	87210	1.12	ng	99
25) Benzo(b)fluoranthene	21.54	252	85644	1.00	ng	98
26) Benzo(k)fluoranthene	21.57	252	102397	1.06	ng	98
27) Benzo(a)pyrene	22.00	252	87397	1.10	ng	99
28) Dibenzo(a,h)anthracene	23.64	278	81204	1.11	ng	99
29) Benzo(a,h,i)perylene	24.24	276	91482	1.05	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021916\
Data File : BE091413.D
Acq On : 19 Feb 2016 13:19
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Quant Time: Feb 19 22:37:56 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 00:27:53 2016
Response via : Initial Calibration

