

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021916\
 Data File : BE091413.D
 Acq On : 19 Feb 2016 13:19
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 19 22:37:56 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2 S	2-Fluorophenol	0.896	1.038	-15.8	127	-0.03
3 S	Phenol-d6	1.352	1.265	6.4	95	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.02
5 S	Nitrobenzene-d5	0.241	0.293	-21.6	139	-0.02
6	Naphthalene	1.128	1.057	6.3	106	-0.02
7	2-Methylnaphthalene	0.609	0.649	-6.6	120	-0.02
8 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
9 S	2,4,6-Tribromophenol	0.153	0.172	-12.4	145	-0.01
10 S	2-Fluorobiphenyl	1.301	1.245	4.3	109	-0.01
11	Acenaphthylene	7.133	7.778	-9.0	130	0.00
12	Acenaphthene	1.373	1.280	6.8	109	0.00
13	Fluorene	1.772	1.721	2.9	110	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
15	Phenanthrene	1.313	1.246	5.1	106	0.00
16	Anthracene	1.100	1.154	-4.9	115	-0.01
17	Fluoranthene	1.203	1.281	-6.5	121	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
19	Pyrene	1.083	1.186	-9.5	121	0.00
20 S	Terphenyl-d14	0.740	0.787	-6.4	114	0.00
21	Benzo(a)anthracene	0.926	1.071	-15.7	135	0.00
22	Chrysene	1.371	1.303	5.0	107	0.00
23	Indeno(1,2,3-cd)pyrene	0.929	1.132	-21.9	131	0.00
24 I	Perylene-d12	1.000	1.000	0.0	106	0.00
25	Benzo(b)fluoranthene	1.048	1.088	-3.8	120	0.00
26	Benzo(k)fluoranthene	1.187	1.301	-9.6	130	0.00
27 C	Benzo(a)pyrene	0.939	1.110	-18.2	139	0.00
28	Dibenzo(a,h)anthracene	0.827	1.032	-24.8	146	0.00
29	Benzo(g,h,i)perylene	1.055	1.162	-10.1	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0