

Data Path : Z:\HPCHEM1\BNA E\DATA\BE021916\
 Data File : BE091413.D
 Acq On : 19 Feb 2016 13:19
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 19 22:37:56 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	-0.01
2 S	2-Fluorophenol	1.000	1.072	-7.2	127	-0.03
3 S	Phenol-d6	1.000	0.940	6.0	95	-0.02
4 I	Naphthalene-d8	5.000	5.000	0.0	100	-0.02
5 S	Nitrobenzene-d5	1.000	1.113	-11.3	139	-0.02
6	Naphthalene	1.000	0.938	6.2	106	-0.02
7	2-Methylnaphthalene	1.000	1.028	-2.8	120	-0.02
8 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
9 S	2,4,6-Tribromophenol	1.000	1.013	-1.3	145	-0.01
10 S	2-Fluorobiphenyl	1.000	0.957	4.3	109	-0.01
11	Acenaphthylene	1.000	1.039	-3.9	130	0.00
12	Acenaphthene	1.000	0.933	6.7	109	0.00
13	Fluorene	1.000	0.971	2.9	110	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	99	-0.01
15	Phenanthrene	1.000	0.949	5.1	106	0.00
16	Anthracene	1.000	1.049	-4.9	115	-0.01
17	Fluoranthene	1.000	1.026	-2.6	121	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	98	0.00
19	Pyrene	1.000	1.094	-9.4	121	0.00
20 S	Terphenyl-d14	1.000	1.065	-6.5	114	0.00
21	Benzo(a)anthracene	1.000	1.098	-9.8	135	0.00
22	Chrysene	1.000	0.995	0.5	107	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	1.125	-12.5	131	0.00
24 I	Perylene-d12	5.000	5.000	0.0	106	0.00
25	Benzo(b)fluoranthene	1.000	0.996	0.4	120	0.00
26	Benzo(k)fluoranthene	1.000	1.056	-5.6	130	0.00
27 C	Benzo(a)pyrene	1.000	1.102	-10.2	139	0.00
28	Dibenzo(a,h)anthracene	1.000	1.110	-11.0	146	0.00
29	Benzo(g,h,i)perylene	1.000	1.051	-5.1	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0