

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022017\
 Data File : BE092750.D
 Acq On : 20 Feb 2017 16:36
 Operator : SJ/MA
 Sample : I1899-02 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SS050SS218(0-6)-170216

Quant Time: Feb 20 17:57:01 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.10	152	11540	5.00	ng	-0.02
7) Naphthalene-d8	10.89	136	54435	5.00	ng	0.00
12) Acenaphthene-d10	14.76	164	32571	5.00	ng	0.00
18) Phenanthrene-d10	17.51	188	71689	5.00	ng	0.00
24) Chrysene-d12	21.68	240	66315	5.00	ng	-0.02
31) Perylene-d12	24.02	264	73895	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.61	112	4596	2.05	ng	-0.02
5) Phenol-d6	7.29	99	5767	2.09	ng	-0.02
8) Nitrobenzene-d5	9.29	82	2836	0.84	ng	-0.02
13) 2,4,6-Tribromophenol	16.25	330	1335	1.93	ng	-0.03
14) 2-Fluorobiphenyl	13.38	172	6544	0.82	ng	-0.02
26) Terphenyl-d14	20.12	244	7404	0.87	ng	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.54	88	8	Below Cal		# 19
3) n-Nitrosodimethylamine	3.69	42	21	0.22	ng	# 23
6) bis(2-Chloroethyl)ether	7.43	93	105	0.03	ng	# 15
9) Naphthalene	10.93	128	256	0.02	ng	# 31
15) Acenaphthylene	14.47	152	17239	0.35	ng	100
16) Acenaphthene	14.82	154	1389	0.19	ng	90
17) Fluorene	15.81	166	1052	0.11	ng	# 97
20) Pentachlorophenol	17.19	266	21	0.28	ng	# 70
21) Phenanthrene	17.56	178	32403	1.90	ng	95
22) Anthracene	17.56	178	32403	2.07	ng	98
23) Fluoranthene	19.54	202	688815	9.43	ng	99
25) Pyrene	19.92	202	574099	9.09	ng	# 93
27) Benzo(a)anthracene	21.72	228	637392	9.91	ng	# 81
28) Chrysene	21.72	228	637392	8.36	ng	# 64
29) Bis(2-ethylhexyl)phthalate	21.59	149	45396	5.11	ng	# 72
30) Indeno(1,2,3-cd)pyrene	26.46	276	240563	3.05	ng	97
32) Benzo(b)fluoranthene	23.30	252	169556	8.71	ng	97
33) Benzo(k)fluoranthene	23.30	252	169556	8.20	ng	96
34) Benzo(a)pyrene	23.91	252	79176	4.17	ng	96
35) Dibenzo(a,h)anthracene	26.48	278	13591	0.73	ng	# 94
36) Benzo(g,h,i)perylene	27.23	276	232018	3.00	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022017\
Data File : BE092750.D
Acq On : 20 Feb 2017 16:36
Operator : SJ/MA
Sample : I1899-02 5X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SS050SS218(0-6)-170216

Quant Time: Feb 20 17:57:01 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE021517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 15 17:05:20 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.10 min Scan# 573

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:152 Resp: 11540

Ion Ratio Lower Upper

152 100

150 187.0 106.0 159.0#




115 96.7 31.2 46.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: Below Cal

RT: 3.54 min Scan# 82

Delta R.T. 0.20 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion: 88 Resp: 8

Ion Ratio Lower Upper

88 100

58 0.0 63.6 95.4#



43 0.0 28.0 42.0#




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.22 ng

RT: 3.69 min Scan# 95

Delta R.T. -0.05 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216




Tot Ion:	42	Resp:	21
Ion	Ratio	Lower	Upper
42	100		
74	23.8	86.0	129.0#
44	0.0	5.1	7.7#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 2.05 ng


RT: 5.61 min Scan# 353

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36



Tgt Ion:	112	Resp:	4596
Ion	Ratio	Lower	Upper
112	100		
64	67.9	53.5	80.3
63	36.2	27.9	41.9




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 2.09 ng

RT: 7.29 min Scan# 538

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion: 99 Resp: 5767

Ion Ratio Lower Upper

99 100

42 0.0 16.0 24.0#




71 0.0 30.6 45.8#



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.03 ng

RT: 7.43 min Scan# 544

Delta R.T. -0.11 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion: 93 Resp: 105

Ion Ratio Lower Upper

93 100

63 0.0 71.6 107.4#



95 0.0 24.0 36.0#




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.89 min Scan# 705

Delta R.T. -0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216

Tot Ion:136 Resp: 54435


Ion Ratio Lower Upper

136 100

137 11.9 8.2 12.4

54 5.1 9.6 14.4#

68 4.0 6.7 10.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.84 ng

RT: 9.29 min Scan# 625

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion: 82 Resp: 2836

Ion Ratio Lower Upper

82 100

128 33.2 39.0 58.4#



54 63.1 44.8 67.2




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.02 ng

RT: 10.93 min Scan# 707

Delta R.T. -0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:128 Resp: 256

Ion Ratio Lower Upper

128 100

129 42.9 11.2 16.8#

127 42.2 11.6 17.4#




Abundance Ion 128.00 (127.70 to 128.70): E


Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E

10.93



Time-->



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.76 min Scan# 977

Delta R.T. 0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion:164 Resp: 32571

Ion Ratio Lower Upper

164 100

162 89.3 71.4 107.0

160 35.3 27.4 41.2




Abundance Ion 164.00 (163.70 to 164.70): E


Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E

14.76







Time-->



#13
 2,4,6-Tribromophenol
 Concen: 1.93 ng
 RT: 16.25 min Scan# 1083
 Delta R.T. -0.03 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36




Instrument :
 BNA_E
 ClientSampleId :
 SS050SS218(0-6)-170216


Tot Ion:330 Resp: 1335
 Ion Ratio Lower Upper
 330 100
 332 97.2 75.4 113.0
 141 44.3 31.0 46.6



#14
 2-Fluorobiphenyl
 Concen: 0.82 ng
 RT: 13.38 min Scan# 879
 Delta R.T. -0.02 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36

Tgt Ion:172 Resp: 6544
 Ion Ratio Lower Upper
 172 100
 171 37.9 30.1 45.1
 170 27.4 21.8 32.6





#15

Acenaphthylene

Concen: 0.35 ng

RT: 14.47 min Scan# 960

Delta R.T. 0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216





Tot Ion:152 Resp: 17239

Ion Ratio Lower Upper

152 100

151 4.8 3.9 5.9

153 13.0 10.4 15.6



#16

Acenaphthene

Concen: 0.19 ng

RT: 14.82 min Scan# 980

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36




Tgt Ion:154 Resp: 1389


Ion Ratio Lower Upper

154 100

153 463.9 350.8 526.2

152 227.8 171.1 256.7





#17

Fluorene

Concen: 0.11 ng

RT: 15.81 min Scan# 1045

Delta R.T. -0.01 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216





Tot Ion:166 Resp: 1052

Ion Ratio Lower Upper

166 100

165 99.7 78.2 117.4

167 18.3 11.0 16.4#



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.51 min Scan# 1151

Delta R.T. 0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36




Tgt Ion:188 Resp: 71689


Ion Ratio Lower Upper

188 100

94 6.2 3.2 4.8#

80 5.5 2.6 3.8#





#20

Pentachlorophenol

Concen: 0.28 ng

RT: 17.19 min Scan# 1137

Delta R.T. -0.11 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:266 Resp: 21

Ion Ratio Lower Upper

266 100

268 100.0 62.5 93.7#





264 100.0 57.2 85.8#



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#21

Phenanthrene

Concen: 1.90 ng

RT: 17.56 min Scan# 1153

Delta R.T. 0.00 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion:178 Resp: 32403

Ion Ratio Lower Upper

178 100

176 19.2 15.8 23.8



179 16.2 16.0 24.0




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#22

Anthracene

Concen: 2.07 ng

RT: 17.56 min Scan# 1153

Delta R.T. -0.09 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:178 Resp: 32403

Ion Ratio Lower Upper

178 100

176 19.2 15.0 22.4




179 16.2 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#23

Fluoranthene

Concen: 9.43 ng

RT: 19.54 min Scan# 1262

Delta R.T. -0.02 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion:202 Resp: 688815

Ion Ratio Lower Upper

202 100

101 2.9 2.2 3.4



203 16.7 13.7 20.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E









#24
Chrysene-d12
Concen: 5.00 ng
RT: 21.68 min Scan# 1372
Delta R.T. -0.02 min
Lab File: BE092750.D
Acq: 20 Feb 2017 16:36




Instrument :
BNA_E
ClientSampleId :
SS050SS218(0-6)-170216


Tot Ion:240 Resp: 66315
Ion Ratio Lower Upper
240 100
120 22.0 2.6 4.0#
236 36.4 24.6 37.0



#25
Pyrene
Concen: 9.09 ng
RT: 19.92 min Scan# 1284
Delta R.T. 0.00 min
Lab File: BE092750.D
Acq: 20 Feb 2017 16:36

Tgt Ion:202 Resp: 574099
Ion Ratio Lower Upper
202 100
200 5.4 4.2 6.4
203 21.2 13.9 20.9#









#26
 Terphenyl-d14
 Concen: 0.87 ng
 RT: 20.12 min Scan# 1296
 Delta R.T. -0.02 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36




Instrument :
 BNA_E
 ClientSampleId :
 SS050SS218(0-6)-170216


Tot Ion:244 Resp: 7404
 Ion Ratio Lower Upper
 244 100
 212 9.1 6.7 10.1
 122 5.6 3.6 5.4#



#27
 Benzo(a)anthracene
 Concen: 9.91 ng
 RT: 21.72 min Scan# 1374
 Delta R.T. 0.05 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36

Tgt Ion:228 Resp: 637392
 Ion Ratio Lower Upper
 228 100
 226 20.8 23.6 35.4#
 229 26.4 13.3 19.9#









#28
 Chrysene
 Concen: 8.36 ng
 RT: 21.72 min Scan# 1374
 Delta R.T. 0.00 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36




Instrument :
 BNA_E
 ClientSampleId :
 SS050SS218(0-6)-170216


Tot Ion:228 Resp: 637392
 Ion Ratio Lower Upper
 228 100
 226 20.8 36.3 54.5#
 229 26.4 10.6 16.0#



#29
 Bis(2-ethylhexyl)phthalate
 Concen: 5.11 ng
 RT: 21.59 min Scan# 1368
 Delta R.T. 0.00 min
 Lab File: BE092750.D
 Acq: 20 Feb 2017 16:36

Tgt Ion:149 Resp: 45396
 Ion Ratio Lower Upper
 149 100
 167 26.0 16.0 24.0#
 279 0.0 16.0 24.0#





#30

Indeno(1,2,3-cd)pyrene

Concen: 3.05 ng

RT: 26.46 min Scan# 1708

Delta R.T. -0.03 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:276 Resp: 240563

Ion Ratio Lower Upper




276 100

138 23.0 20.3 30.5

277 23.5 19.7 29.5



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E



#31

Perylene-d12

Concen: 5.00 ng

RT: 24.02 min Scan# 1538

Delta R.T. -0.01 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion:264 Resp: 73895

Ion Ratio Lower Upper



264 100


260 26.6 20.1 30.1

265 21.5 17.9 26.9



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#32
Benzo(b)fluoranthene
Concen: 8.71 ng
RT: 23.30 min Scan# 1476
Delta R.T. -0.01 min
Lab File: BE092750.D
Acq: 20 Feb 2017 16:36




Instrument :
BNA_E
ClientSampleId :
SS050SS218(0-6)-170216


Tot Ion:252 Resp: 169556
Ion Ratio Lower Upper
252 100
253 22.3 18.7 28.1
125 11.6 10.5 15.7



#33
Benzo(k)fluoranthene
Concen: 8.20 ng
RT: 23.30 min Scan# 1476
Delta R.T. -0.06 min
Lab File: BE092750.D
Acq: 20 Feb 2017 16:36

Tgt Ion:252 Resp: 169556
Ion Ratio Lower Upper
252 100
253 22.3 20.3 30.5
125 11.6 9.6 14.4





#34

Benzo(a)pyrene

Concen: 4.17 ng

RT: 23.91 min Scan# 1529

Delta R.T. -0.01 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216


Tot Ion:252 Resp: 79176

Ion Ratio Lower Upper




252 100

253 23.0 19.0 28.6

125 12.0 11.8 17.8



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#35

Dibenzo(a,h)anthracene

Concen: 0.73 ng

RT: 26.48 min Scan# 1709

Delta R.T. -0.03 min

Lab File: BE092750.D

Acq: 20 Feb 2017 16:36


Tgt Ion:278 Resp: 13591

Ion Ratio Lower Upper



278 100


139 21.1 13.8 20.8#

279 28.8 21.4 32.2



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#36

Benzo(a,h,i)perylene

Concen: 3.00 ng

RT: 27.23 min Scan# 1753

Delta R.T. -0.02 min

Lab File: BE092750.D


Acq: 20 Feb 2017 16:36

Instrument :

BNA_E

ClientSampleId :

SS050SS218(0-6)-170216




Tot Ion: 276 Resp: 232018

Ion Ratio Lower Upper

276 100

277 25.3 19.8 29.8

138 22.5 19.8 29.6



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

