

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022017\
 Data File : BE092746.D
 Acq On : 20 Feb 2017 14:07
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 10:16:00 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.710	0.642	9.6	89	0.00
3	n-Nitrosodimethylamine	0.740	0.479	35.3#	72	0.02
4 S	2-Fluorophenol	0.973	0.778	20.0	81	0.01
5 S	Phenol-d6	1.194	1.037	13.1	88	0.00
6	bis(2-Chloroethyl)ether	1.321	1.185	10.3	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.279	0.261	6.5	95	0.00
9	Naphthalene	1.082	1.126	-4.1	101	0.00
10	Hexachlorobutadiene	0.159	0.165	-3.8	98	0.00
11	2-Methylnaphthalene	0.680	0.664	2.4	98	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	0.106	0.093	12.3	90	0.01
14 S	2-Fluorobiphenyl	1.219	1.245	-2.1	101	0.00
15	Acenaphthylene	7.544	7.504	0.5	100	0.00
16	Acenaphthene	1.096	1.110	-1.3	100	-0.02
17	Fluorene	1.407	1.389	1.3	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19	Hexachlorobenzene	0.193	0.207	-7.3	98	0.00
20	Pentachlorophenol	0.042	0.016	61.9#	46#	0.12
21	Phenanthrene	1.188	1.203	-1.3	97	0.00
22	Anthracene	1.094	1.053	3.7	99	0.02
23	Fluoranthene	5.096	4.946	2.9	98	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
25	Pyrene	4.761	4.811	-1.1	96	0.00
26 S	Terphenyl-d14	0.638	0.638	0.0	95	0.00
27	Benzo(a)anthracene	4.849	4.977	-2.6	97	0.00
28	Chrysene	5.752	6.377	-10.9	103	0.00
29	Bis(2-ethylhexyl)phthalate	0.674	0.545	19.1	87	0.00
30	Indeno(1,2,3-cd)pyrene	6.486	5.701	12.1	81	0.00
31 I	Perylene-d12	1.000	1.000	0.0	96	0.00
32	Benzo(b)fluoranthene	1.414	1.270	10.2	89	0.00
33	Benzo(k)fluoranthene	1.495	1.271	15.0	89	0.00
34 C	Benzo(a)pyrene	1.285	1.119	12.9	89	-0.01
35	Dibenzo(a,h)anthracene	1.355	1.118	17.5	80	0.00
36	Benzo(g,h,i)perylene	5.728	4.896	14.5	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0