

Data Path : Z:\HPCHEM1\BNA_E\Data\BE022017\
 Data File : BE092753.D
 Acq On : 20 Feb 2017 22:52
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 00:54:24 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE021517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 17:05:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	185	0.00
2	1,4-Dioxane	0.400	0.374	6.5	172	0.00
3	n-Nitrosodimethylamine	0.400	0.499	-24.7	199	-0.01
4 S	2-Fluorophenol	0.400	0.483	-20.7	232	-0.02
5 S	Phenol-d6	0.400	0.588	-47.0#	281	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	218	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	195	-0.02
8 S	Nitrobenzene-d5	0.400	0.456	-14.0	234	-0.02
9	Naphthalene	0.400	0.408	-2.0	197	0.00
10	Hexachlorobutadiene	0.400	0.432	-8.0	202	0.00
11	2-Methylnaphthalene	0.400	0.400	0.0	200	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	201	0.00
13 S	2,4,6-Tribromophenol	0.400	0.518	-29.5#	266	-0.02
14 S	2-Fluorobiphenyl	0.400	0.426	-6.5	211	-0.02
15	Acenaphthylene	0.400	0.397	0.8	201	0.00
16	Acenaphthene	0.400	0.411	-2.7	205	-0.02
17	Fluorene	0.400	0.397	0.8	201	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	202	0.00
19	Hexachlorobenzene	0.400	0.383	4.3	176	0.00
20	Pentachlorophenol	0.400	0.553	-38.3#	246	-0.09
21	Phenanthrene	0.400	0.368	8.0	176	0.00
22	Anthracene	0.400	0.399	0.3	206	0.00
23	Fluoranthene	0.400	0.378	5.5	190	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	154	0.00
25	Pyrene	0.400	0.487	-21.7	193	0.00
26 S	Terphenyl-d14	0.400	0.517	-29.3#	205	-0.02
27	Benzo(a)anthracene	0.400	0.446	-11.5	175	0.00
28	Chrysene	0.400	0.415	-3.7	159	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.702	-75.5#	288	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.443	-10.7	146	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	158	-0.01
32	Benzo(b)fluoranthene	0.400	0.422	-5.5	148	-0.01
33	Benzo(k)fluoranthene	0.400	0.398	0.5	160	-0.01
34 C	Benzo(a)pyrene	0.400	0.390	2.5	163	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.389	2.8	143	-0.02
36	Benzo(g,h,i)perylene	0.400	0.383	4.3	138	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0