

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\Be022119\
 Data File : BE098746.D
 Acq On : 21 Feb 2019 13:10
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 13:38:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.739	0.682	7.7	110	0.00
3	n-Nitrosodimethylamine	0.877	0.888	-1.3	111	0.00
4 S	2-Fluorophenol	1.182	1.226	-3.7	115	0.00
5 S	Phenol-d6	1.737	1.813	-4.4	124	0.00
6	bis(2-Chloroethyl)ether	1.183	1.248	-5.5	130	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.389	0.399	-2.6	119	0.00
9	Naphthalene	4.584	4.502	1.8	118	0.00
10	Hexachlorobutadiene	0.307	0.289	5.9	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.757	1.0	116	0.00
12	2-Methylnaphthalene	0.761	0.754	0.9	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.177	0.184	-4.0	126	0.00
15 S	2-Fluorobiphenyl	1.714	1.559	9.0	114	0.00
16	Acenaphthylene	2.052	1.932	5.8	119	0.00
17	Acenaphthene	1.212	1.153	4.9	116	0.00
18	Fluorene	1.656	1.596	3.6	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.236	0.228	3.4	119	-0.01
21	Hexachlorobenzene	0.278	0.264	5.0	114	0.00
22	Pentachlorophenol	0.071	0.063	11.3	116	0.00
23	Phenanthrene	1.101	1.070	2.8	119	0.00
24	Anthracene	0.919	0.905	1.5	123	0.00
25 SURR	Fluoranthene-d10	5.713	5.621	1.6	120	0.00
26	Fluoranthene	1.417	1.386	2.2	120	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
28	Pyrene	1.274	1.303	-2.3	122	0.00
29 S	Terphenyl-d14	0.998	0.984	1.4	116	0.00
30	Benzo(a)anthracene	1.207	1.250	-3.6	129	0.00
31	Chrysene	1.308	1.254	4.1	115	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	2.971	-32.8#	162#	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.238	9.6	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.294	1.265	2.2	117	0.00
36	Benzo(k)fluoranthene	1.383	1.303	5.8	116	0.00
37 C	Benzo(a)pyrene	1.237	1.204	2.7	120	0.00
38	Dibenzo(a,h)anthracene	1.146	1.005	12.3	114	0.00
39	Benzo(g,h,i)perylene	1.258	1.136	9.7	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0