

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\Be022119\
 Data File : BE098753.D
 Acq On : 21 Feb 2019 18:05
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 22:12:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE012419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 24 14:44:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	140	0.00
2	1,4-Dioxane	0.739	0.701	5.1	141	0.00
3	n-Nitrosodimethylamine	0.877	1.098	-25.2#	171#	0.00
4 S	2-Fluorophenol	1.182	1.191	-0.8	139	0.00
5 S	Phenol-d6	1.737	1.842	-6.0	156#	0.00
6	bis(2-Chloroethyl)ether	1.183	1.387	-17.2	180#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	166#	0.00
8 S	Nitrobenzene-d5	0.389	0.395	-1.5	169#	-0.01
9	Naphthalene	4.584	4.509	1.6	168#	0.00
10	Hexachlorobutadiene	0.307	0.290	5.5	160#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.765	0.765	0.0	168#	0.00
12	2-Methylnaphthalene	0.761	0.767	-0.8	170#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	177#	0.00
14 S	2,4,6-Tribromophenol	0.177	0.194	-9.6	189#	-0.01
15 S	2-Fluorobiphenyl	1.714	1.707	0.4	178#	-0.02
16	Acenaphthylene	2.052	2.036	0.8	178#	0.00
17	Acenaphthene	1.212	1.202	0.8	171#	-0.02
18	Fluorene	1.656	1.661	-0.3	179#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	175#	-0.01
20	4-Bromophenyl-phenylether	0.236	0.230	2.5	178#	-0.01
21	Hexachlorobenzene	0.278	0.254	8.6	163#	0.00
22	Pentachlorophenol	0.071	0.056	21.1	154#	0.00
23	Phenanthrene	1.101	1.095	0.5	181#	0.00
24	Anthracene	0.919	0.917	0.2	185#	0.00
25 SURR	Fluoranthene-d10	5.713	5.566	2.6	176#	0.00
26	Fluoranthene	1.417	1.405	0.8	180#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	174#	-0.01
28	Pyrene	1.274	1.276	-0.2	180#	0.00
29 S	Terphenyl-d14	0.998	0.991	0.7	176#	0.00
30	Benzo(a)anthracene	1.207	1.279	-6.0	198#	0.00
31	Chrysene	1.308	1.260	3.7	174#	0.00
32	Bis(2-ethylhexyl)phthalate	2.238	3.196	-42.8#	261#	0.00
33	Indeno(1,2,3-cd)pyrene	1.370	1.315	4.0	182#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	182#	0.00
35	Benzo(b)fluoranthene	1.294	1.299	-0.4	185#	0.00
36	Benzo(k)fluoranthene	1.383	1.329	3.9	181#	0.00
37 C	Benzo(a)pyrene	1.237	1.199	3.1	183#	0.00
38	Dibenzo(a,h)anthracene	1.146	1.066	7.0	185#	0.00
39	Benzo(g,h,i)perylene	1.258	1.164	7.5	176#	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0