

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
 Data File : BE089654.D
 Acq On : 23 Feb 2015 23:29
 Operator : TP/JJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Quant Time: Feb 24 03:18:16 2015

Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Feb 24 03:06:59 2015

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.34	152	41455	5.00	ng	0.00
4) Naphthalene-d8	12.24	136	176304	5.00	ng	0.00
8) Acenaphthene-d10	16.42	164	74055	5.00	ng	0.00
13) Phenanthrene-d10	19.77	188	134055	5.00	ng	0.00
19) Chrysene-d12	23.66	240	116610	5.00	ng	0.00
25) Perylene-d12	25.34	264	92624	5.00	ng	0.00

System Monitoring Compounds

5) Nitrobenzene-d5	10.60	82	48432	6.04	ng	0.00
9) 2-Fluorobiphenyl	14.88	172	82923	4.29	ng	0.00
21) Terphenyl-d14	22.30	244	73511	3.92	ng	0.00

Target Compounds

						Ovalue
2) 1,4-Dioxane	2.69	88	25888	3.77	ng	99
3) n-Nitrosodimethylamine	3.32	42	35909	4.91	ng	98
6) Naphthalene	12.29	128	170722	4.26	ng	100
7) 2-Methylnaphthalene	13.95	142	101467	4.50	ng	96
10) Acenaphthylene	16.07	152	584222	4.76	ng	100
11) Acenaphthene	16.49	154	86353	4.40	ng	99
12) Fluorene	17.74	166	99131	4.83	ng	100
14) Hexachlorobenzene	19.00	284	17676	4.27	ng	99
15) Pentachlorophenol	19.43	266	3799	3.69	ng	94
16) Phenanthrene	19.81	178	144431	4.30	ng	100
17) Anthracene	19.91	178	138464	5.10	ng	99
18) Fluoranthene	21.72	202	139425	5.72	ng	100
20) Pyrene	22.04	202	151603	4.14	ng	99
22) Benzo(a)anthracene	23.64	228	459546	5.37	ng	96
23) Chrysene	23.68	228	516361	4.09	ng	95
24) Indeno(1,2,3-cd)pyrene	26.74	276	347354	7.30	ng	# 91
26) Benzo(b)fluoranthene	24.91	252	76153	8.59	ng	97
27) Benzo(k)fluoranthene	24.94	252	128536	5.53	ng	98
28) Benzo(a)pyrene	25.27	252	77200	7.55	ng	97
29) Dibenzo(a,h)anthracene	26.78	278	67169	7.83	ng	94
30) Benzo(g,h,i)perylene	27.16	276	311226	5.85	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
Data File : BE089654.D
Acq On : 23 Feb 2015 23:29
Operator : TP/JJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Feb 24 03:18:16 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 24 03:06:59 2015
Response via : Initial Calibration

