

Data Path : Z:\HPCHEM1\BNA E\DATA\BE022315\
 Data File : BE089656.D
 Acq On : 24 Feb 2015 00:47
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Quant Time: Feb 24 04:03:36 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE022315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 24 03:46:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.642	0.649	-1.1	97	0.00
3	n-Nitrosodimethylamine	0.745	0.894	-20.0	106	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
5 S	Nitrobenzene-d5	0.291	0.263	9.6	97	0.00
6	Naphthalene	1.021	1.024	-0.3	104	0.00
7	2-Methylnaphthalene	0.575	0.591	-2.8	104	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
9 S	2-Fluorobiphenyl	1.355	1.212	10.6	100	0.00
10	Acenaphthylene	7.371	7.648	-3.8	105	-0.01
11	Acenaphthene	1.194	1.208	-1.2	105	0.00
12	Fluorene	1.311	1.353	-3.2	104	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
14	Hexachlorobenzene	0.136	0.141	-3.7	94	0.00
15	Pentachlorophenol	0.020	0.015	25.0#	95	0.00
16	Phenanthrene	1.103	1.119	-1.5	90	0.00
17	Anthracene	0.931	0.961	-3.2	90	0.00
18	Fluoranthene	0.917	0.963	-5.0	102	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
20	Pvrene	1.256	1.302	-3.7	103	0.00
21 S	Terphenyl-d14	0.714	0.775	-8.5	78	0.00
22	Benzo(a)anthracene	3.489	3.460	0.8	101	0.00
23	Chrysene	4.726	4.808	-1.7	104	0.00
24	Indeno(1,2,3-cd)pyrene	1.955	1.822	6.8	95	0.00
25 I	Perylene-d12	1.000	1.000	0.0	116	0.00
26	Benzo(b)fluoranthene	0.525	0.455	13.3	102	0.00
27	Benzo(k)fluoranthene	1.190	1.383	-16.2	100	0.00
28 C	Benzo(a)pyrene	0.565	0.532	5.8	98	0.00
29	Dibenzo(a,h)anthracene	0.466	0.440	5.6	96	0.00
30	Benzo(g,h,i)perylene	2.566	2.587	-0.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0